



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118135

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	01/10/2022	10:45
2	N1042-01		1	8	100.00	>212.0	>212.0	01/10/2022	11:10
3	N1056-06		1	7	40.00	104.0	103.7	01/10/2022	11:35
4	N1056-06DUP		1	8	41.00	105.8	105.5	01/10/2022	11:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118135

WorkList Name : fp11022 WorkList ID : 155879 Department : Wet-Chemistry Date : 01-10-2022 10:24:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2022	Solid	N1042-01	Flash Point	Cool 4 deg C	PSEG03	I11	TRE-1268	01/05/2022	1010A
01/11/2022	Solid	N1056-06	Flash Point	Cool 4 deg C	PSEG03	I21	1917-OIL	01/06/2022	1010A

Date/Time 01/10/22 10:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 01/10/22 12:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)