



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118136

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	01/10/2022	10:45
2	N1050-03		1	12	100.00	>212.0	>212.0	01/10/2022	12:25
3	N1050-03DUP		1	12	100.00	>212.0	>212.0	01/10/2022	12:50
4	N1056-07		1	15	100.00	>212.0	>212.0	01/10/2022	13:15
5	N1056-08		1	13	100.00	>212.0	>212.0	01/10/2022	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB118136

WorkList Name : fp1722 WorkList ID : 155825 Department : Wet-Chemistry Date : 01-07-2022 09:28:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/13/2022	Water	N1050-03	Flash Point	Cool 4 deg C	PSEG03	I21	P1660	01/06/2022	1010B
01/11/2022	Water	N1056-07	Flash Point	Cool 4 deg C	PSEG03	I21	COMP-3	01/06/2022	1010B
01/11/2022	Water	N1056-08	Flash Point	Cool 4 deg C	PSEG03	I21	1917-WATER	01/06/2022	1010B

Date/Time 01/07/22 09:55
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: RJ (1721)ab

Date/Time 01/10/22 13:45
Raw Sample Received by: RJ & Ext-lab
Raw Sample Relinquished by: KS(WC)