


LB118142

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.025	0.025	68	0.000	68					1
0.2	0.200	0.196	-0.004	915	0.0 %	915					-2
2.5	2.500	2.463	-0.037	12184	0.0 %	12184					-1
5.0	5.000	5.023	0.023	24912	0.0 %	24912					0
7.5	7.500	7.485	-0.015	37148	0.0 %	37148					0
10.0	10.000	10.009	0.009	49697	0.0 %	49697					0

Method: 7471B		Operator: Admin		Date of Analysis: 10 Jan 2022 14:27:27					
Sample ID	Extended ID	µ Abs	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	7471B	PPB	10 Jan 2022	14:33:31	S Std
0.2 - 1	50.2	915	-	0.2000	7471B	PPB	10 Jan 2022	14:35:48	S Std
2.5 - 1	52.5	12184	-	2.5000	7471B	PPB	10 Jan 2022	14:38:06	S Std
5.0 - 1	55	24912	-	5.0000	7471B	PPB	10 Jan 2022	14:40:23	S Std
7.5 - 1	57.5	37148	-	7.5000	7471B	PPB	10 Jan 2022	14:47:51	S Std
10.0 - 1	510	49697	-	10.0000	7471B	PPB	10 Jan 2022	14:52:42	S Std
ICV38 - 1	ICV38	20741	4.1840	-	7471B	PPB	10 Jan 2022	14:55:23	U SMPL
ICB38 - 1	ICB38	-8	0.0098	-	7471B	PPB	10 Jan 2022	14:57:38	U SMPL
CCV67 - 1	CCV67	24624	4.9651	-	7471B	PPB	10 Jan 2022	14:59:53	U SMPL
CCB67 - 1	CCB67	-3	0.0108	-	7471B	PPB	10 Jan 2022	15:02:08	U SMPL
CRA - 1	CRA	901	0.1927	-	7471B	PPB	10 Jan 2022	15:04:24	U SMPL
HighStd - 1	HighStd	50183	10.1069	-	7471B	PPB	10 Jan 2022	15:06:38	U SMPL
ChkStd - 1	ChkStd	36522	7.3587	-	7471B	PPB	10 Jan 2022	15:08:54	U SMPL
PB141945BL - 1	PBS	-97	-0.0081	-	7471B	PPB	10 Jan 2022	15:11:12	U SMPL
PB141945BS - 1	LCSS	19876	4.0100	-	7471B	PPB	10 Jan 2022	15:13:28	U SMPL
N1049-01 - 1	CHRT-20430-27106-26948-24245	240	0.0597	-	7471B	PPB	10 Jan 2022	15:15:44	U SMPL
N1049-03 - 1	7211979	859	0.1842	-	7471B	PPB	10 Jan 2022	15:18:01	U SMPL
N1049-05 - 1	CHRT-24368-26855-26562-27165-20413	918	0.1961	-	7471B	PPB	10 Jan 2022	15:20:18	U SMPL
N1050-01 - 1	VNJ-236	1835	0.3806	-	7471B	PPB	10 Jan 2022	15:22:34	U SMPL
N1052-01 - 1	BALL-1	11570	2.3390	-	7471B	PPB	10 Jan 2022	15:24:50	U SMPL
CCV68 - 1	CCV68	24681	4.9766	-	7471B	PPB	10 Jan 2022	15:27:05	U SMPL
CCB68 - 1	CCB68	18	0.0151	-	7471B	PPB	10 Jan 2022	15:29:20	U SMPL
N1064-01 - 1	HD-02-010622	961	0.2048	-	7471B	PPB	10 Jan 2022	15:31:37	U SMPL
N1067-01 - 1	124035	1835	0.3806	-	7471B	PPB	10 Jan 2022	15:33:52	U SMPL
N1068-01 - 1	TR-05-010622	603	0.1327	-	7471B	PPB	10 Jan 2022	15:36:08	U SMPL
N1070-01 - 1	HR-2-010622	2568	0.5280	-	7471B	PPB	10 Jan 2022	15:38:24	U SMPL
N1070-03 - 1	HR-03-010622	2706	0.5558	-	7471B	PPB	10 Jan 2022	15:40:39	U SMPL
N1071-01 - 1	FM-E6-01-4.5	3650	0.7457	-	7471B	PPB	10 Jan 2022	15:42:56	U SMPL
N1071-02 - 1	FM-E1-01-4.5	8317	1.6846	-	7471B	PPB	10 Jan 2022	15:45:12	U SMPL
N1071-03 - 1	FM-0226-01-4.5	2527	0.5198	-	7471B	PPB	10 Jan 2022	15:47:29	U SMPL
N1071-03DUP - 1	FM-0226-01-4.5DUP	2538	0.5220	-	7471B	PPB	10 Jan 2022	15:49:46	U SMPL
N1071-03MS - 1	FM-0226-01-4.5MS	23739	4.7871	-	7471B	PPB	10 Jan 2022	15:52:01	U SMPL
CCV69 - 1	CCV69	24930	5.0267	-	7471B	PPB	10 Jan 2022	15:54:17	U SMPL
CCB69 - 1	CCB69	5	0.0124	-	7471B	PPB	10 Jan 2022	15:56:32	U SMPL
N1071-03MSD - 1	FM-0226-01-4.5MSD	23085	4.6555	-	7471B	PPB	10 Jan 2022	15:58:48	U SMPL
PB141946BL - 1	PBS	-55	0.0004	-	7471B	PPB	10 Jan 2022	16:01:03	U SMPL
PB141946BS - 1	LCSS	19358	3.9057	-	7471B	PPB	10 Jan 2022	16:03:19	U SMPL
N1065-01 - 1	AMI-1	655	0.1432	-	7471B	PPB	10 Jan 2022	16:05:34	U SMPL
N1065-06 - 1	AMI-2	1655	0.3444	-	7471B	PPB	10 Jan 2022	16:07:51	U SMPL
N1065-10 - 1	AMI-3	1694	0.3522	-	7471B	PPB	10 Jan 2022	16:10:07	U SMPL
N1065-14 - 1	AMI-4	1966	0.4069	-	7471B	PPB	10 Jan 2022	16:12:23	U SMPL
N1065-18 - 1	AMI-5	4386	0.8938	-	7471B	PPB	10 Jan 2022	16:14:40	U SMPL
N1065-22 - 1	AMI-6	502	0.1124	-	7471B	PPB	10 Jan 2022	16:16:57	U SMPL
N1065-26 - 1	AMI-7	1057	0.2241	-	7471B	PPB	10 Jan 2022	16:19:12	U SMPL
CCV70 - 1	CCV70	25168	5.0746	-	7471B	PPB	10 Jan 2022	16:21:28	U SMPL
CCB70 - 1	CCB70	-13	0.0088	-	7471B	PPB	10 Jan 2022	16:23:44	U SMPL
N1065-30 - 1	AMI-8	1714	0.3562	-	7471B	PPB	10 Jan 2022	16:26:00	U SMPL
N1065-34 - 1	AMI-9	2370	0.4882	-	7471B	PPB	10 Jan 2022	16:28:15	U SMPL
N1065-38 - 1	AMI-10	835	0.1794	-	7471B	PPB	10 Jan 2022	16:30:31	U SMPL
N1065-42 - 1	AMI-11	4281	0.8727	-	7471B	PPB	10 Jan 2022	16:32:46	U SMPL
N1065-42DUP - 1	AMI-11DUP	4207	0.8578	-	7471B	PPB	10 Jan 2022	16:35:02	U SMPL
N1065-42MS - 1	AMI-11MS	26719	5.3866	-	7471B	PPB	10 Jan 2022	16:37:19	U SMPL
N1065-42MSD - 1	AMI-11MSD	26077	5.2574	-	7471B	PPB	10 Jan 2022	16:39:36	U SMPL
CCV71 - 1	CCV71	25326	5.1064	-	7471B	PPB	10 Jan 2022	16:41:53	U SMPL
CCB71 - 1	CCB71	-8	0.0098	-	7471B	PPB	10 Jan 2022	16:44:11	U SMPL
N1071-03LX5 - 1		504	0.1128	-	7471B	PPB	10 Jan 2022	16:46:27	U SMPL
N1071-03A - 1		23643	4.7678	-	7471B	PPB	10 Jan 2022	16:48:43	U SMPL
N1065-42LX5 - 1		1053	0.2233	-	7471B	PPB	10 Jan 2022	16:50:58	U SMPL
N1065-42A - 1		25069	5.0547	-	7471B	PPB	10 Jan 2022	16:53:14	U SMPL
CCV72 - 1	CCV72	25043	5.0494	-	7471B	PPB	10 Jan 2022	16:55:29	U SMPL
CCB72 - 1	CCB72	32	0.0179	-	7471B	PPB	10 Jan 2022	16:57:45	U SMPL