

LB118150


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37637



A= 0.0000e+000

B= 2.6858e-004

C= 3.4876e-003

Rho= 0.9995388

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.0	0.000	0.019	0.019	56	0.000	56					-
0.2	0.200	0.196	-0.004	715	0.0 %	715					-2
2.5	2.500	2.498	-0.002	9288	0.0 %	9288					0
5.0	5.000	5.103	0.103	18989	0.0 %	18989					2
7.5	7.500	7.273	-0.227	27065	0.0 %	27065					-3
10.0	10.000	10.112	0.112	37637	0.0 %	37637					1

LB118150 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jan 2022 08:41:49

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	56	-	0.0007470A	PPB	11 Jan 2022 08:57:52	S	Std
0.2 - 1	50.2	715	-	0.2007470A	PPB	11 Jan 2022 09:00:10	S	Std
2.5 - 1	52.5	9288	-	2.5007470A	PPB	11 Jan 2022 09:02:27	S	Std
5.0 - 1	55	18989	-	5.0007470A	PPB	11 Jan 2022 09:04:44	S	Std
7.5 - 1	57.5	27065	-	7.5007470A	PPB	11 Jan 2022 09:09:32	S	Std
10.0 - 1	51.0	37637	-	10.0007470A	PPB	11 Jan 2022 09:11:47	S	Std
ICV39 - 1	ICV39	14811	3.9814	-7470A	PPB	11 Jan 2022 09:18:02	U	SMPL
ICB39 - 1	ICB39	-17	-0.0011	-7470A	PPB	11 Jan 2022 09:20:16	U	SMPL
CCV73 - 1	CCV73	18126	4.8717	-7470A	PPB	11 Jan 2022 09:30:07	U	SMPL
CCB73 - 1	CCB73	-15	-0.0005	-7470A	PPB	11 Jan 2022 09:32:22	U	SMPL
CRA - 1	CRA	640	0.1754	-7470A	PPB	11 Jan 2022 09:34:37	U	SMPL
HighStd - 1	HighStd	35203	9.4582	-7470A	PPB	11 Jan 2022 09:36:52	U	SMPL
ChkStd - 1	ChkStd	25722	6.9118	-7470A	PPB	11 Jan 2022 09:39:07	U	SMPL
PB141943BL - 1	PBW	-85	-0.0193	-7470A	PPB	11 Jan 2022 09:41:24	U	SMPL
PB141943BS - 1	LCSW	14582	3.9199	-7470A	PPB	11 Jan 2022 09:46:11	U	SMPL
N1065-02 - 1	AMI-1	-17	-0.0011	-7470A	PPB	11 Jan 2022 09:48:26	U	SMPL
N1065-07 - 1	AMI-2	42	0.0148	-7470A	PPB	11 Jan 2022 09:50:43	U	SMPL
N1065-11 - 1	AMI-3	37	0.0134	-7470A	PPB	11 Jan 2022 09:53:00	U	SMPL
N1065-15 - 1	AMI-4	56	0.0185	-7470A	PPB	11 Jan 2022 09:55:17	U	SMPL
N1065-19 - 1	AMI-5	43	0.0150	-7470A	PPB	11 Jan 2022 09:57:32	U	SMPL
CCV74 - 1	CCV74	19062	5.1231	-7470A	PPB	11 Jan 2022 09:59:47	U	SMPL
CCB74 - 1	CCB74	-4	0.0024	-7470A	PPB	11 Jan 2022 10:02:03	U	SMPL
N1065-23 - 1	AMI-6	43	0.0150	-7470A	PPB	11 Jan 2022 10:04:19	U	SMPL
N1065-27 - 1	AMI-7	43	0.0150	-7470A	PPB	11 Jan 2022 10:06:34	U	SMPL
N1065-31 - 1	AMI-8	20	0.0089	-7470A	PPB	11 Jan 2022 10:08:49	U	SMPL
N1065-35 - 1	AMI-9	19	0.0086	-7470A	PPB	11 Jan 2022 10:11:05	U	SMPL
N1065-39 - 1	AMI-10	61	0.0199	-7470A	PPB	11 Jan 2022 10:13:21	U	SMPL
N1065-43 - 1	AMI-11	55	0.0183	-7470A	PPB	11 Jan 2022 10:15:37	U	SMPL
N1065-43DUP - 1	AMI-11DUP	76	0.0239	-7470A	PPB	11 Jan 2022 10:17:53	U	SMPL
N1065-43MS - 1	AMI-11MS	14254	3.8318	-7470A	PPB	11 Jan 2022 10:20:10	U	SMPL
N1065-43MSD - 1	AMI-11MSD	13099	3.5216	-7470A	PPB	11 Jan 2022 10:22:27	U	SMPL
PB141944BL - 1	PBW	13	0.0070	-7470A	PPB	11 Jan 2022 10:24:43	U	SMPL
CCV75 - 1	CCV75	19078	5.1274	-7470A	PPB	11 Jan 2022 10:26:58	U	SMPL
CCB75 - 1	CCB75	-9	0.0011	-7470A	PPB	11 Jan 2022 10:29:13	U	SMPL
PB141944BS - 1	LCSW	12313	3.3105	-7470A	PPB	11 Jan 2022 10:31:29	U	SMPL
N1054-08 - 1	20220019-CAPE-1-MODIFIED-ELUTRIATE	76	0.0239	-7470A	PPB	11 Jan 2022 10:33:45	U	SMPL
N1054-08DUP - 1	20220019-CAPE-1-MODIFIED-ELUTRIATEDUP	70	0.0223	-7470A	PPB	11 Jan 2022 10:36:00	U	SMPL
N1054-08MS - 1	20220019-CAPE-1-MODIFIED-ELUTRIATEMS	12140	3.2640	-7470A	PPB	11 Jan 2022 10:38:16	U	SMPL
N1054-08MSD - 1	20220019-CAPE-1-MODIFIED-ELUTRIATEMSD	12989	3.4920	-7470A	PPB	11 Jan 2022 10:40:31	U	SMPL
N1054-10 - 1	20220020-CAPE-2-MODIFIED-ELUTRIATE	-11	0.0005	-7470A	PPB	11 Jan 2022 10:42:48	U	SMPL
N1054-12 - 1	20220021-CAPE-3-MODIFIED-ELUTRIATE	171	0.0494	-7470A	PPB	11 Jan 2022 10:45:06	U	SMPL
N1054-14 - 1	20220022-CAPE-4-MODIFIED-ELUTRIATE	85	0.0263	-7470A	PPB	11 Jan 2022 10:47:23	U	SMPL
N1054-16 - 1	20220023-CAPE-5-MODIFIED-ELUTRIATE	186	0.0534	-7470A	PPB	11 Jan 2022 10:49:40	U	SMPL
N1054-18 - 1	20220024-CAPE-6-MODIFIED-ELUTRIATE	81	0.0252	-7470A	PPB	11 Jan 2022 10:51:56	U	SMPL
CCV76 - 1	CCV76	18868	5.0710	-7470A	PPB	11 Jan 2022 10:54:11	U	SMPL
CCB76 - 1	CCB76	-35	-0.0059	-7470A	PPB	11 Jan 2022 10:56:27	U	SMPL
PB141915TB - 1	PB141915TB	21	0.0091	-7470A	PPB	11 Jan 2022 10:58:43	U	SMPL
N1065-43LX5 - 1		-8	0.0013	-7470A	PPB	11 Jan 2022 11:00:58	U	SMPL
N1065-43A - 1		14577	3.9185	-7470A	PPB	11 Jan 2022 11:03:14	U	SMPL
N1054-08LX5 - 1		-33	-0.0054	-7470A	PPB	11 Jan 2022 11:05:29	U	SMPL
N1054-08A - 1		14731	3.9599	-7470A	PPB	11 Jan 2022 11:07:45	U	SMPL
CCV77 - 1	CCV77	18924	5.0860	-7470A	PPB	11 Jan 2022 11:10:02	U	SMPL
CCB77 - 1	CCB77	-11	0.0005	-7470A	PPB	11 Jan 2022 11:12:18	U	SMPL