



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118151

Slope : 98.6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/11/2022	08:25
2	CAL2	1	Water	NA	NA	20.3	7.00	01/11/2022	08:26
3	CAL3	1	Water	NA	NA	20.3	10.03	01/11/2022	08:27
4	ICV	1	Water	NA	NA	20.6	7.02	01/11/2022	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	01/11/2022	08:32
6	N1087-01	1	Water	NA	NA	20.2	6.86	01/11/2022	08:35
7	N1087-01DUP	1	Water	NA	NA	20.3	6.87	01/11/2022	08:36
8	CCV2	1	Water	NA	NA	20.4	12.02	01/11/2022	08:39

WB 118151

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1087 WorkList ID : 155888 Department : Wet-Chemistry Date : 01-11-2022 08:19:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2022	Water	N1087-01	pH	Cool 4 deg C	PSEG03	I11	FMI-114	01/10/2022	9040C

Date/Time 01-11-22 08:20
Raw Sample Received by: SP(WC)
Raw Sample Relinquished by: SP(WC)

Date/Time 01-11-22 10:00
Raw Sample Received by: SP(WC)
Raw Sample Relinquished by: SP(WC)