

LB118156


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47850



A= 0.0000e+000

B= 2.0937e-004

C= -1.0085e-002

Rho= 0.9999736

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	-0.003	-0.003	33	0.000	33					-
0.2	0.200	0.197	-0.003	987	0.0 %	987					-2
2.5	2.500	2.540	0.040	12180	0.0 %	12180					2
5.0	5.000	4.950	-0.050	23688	0.0 %	23688					-1
7.5	7.500	7.509	0.009	35911	0.0 %	35911					0
10.0	10.000	10.008	0.008	47850	0.0 %	47850					0

LB118156 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 11 Jan 2022 13:34:10				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	33	-	0.00007471B	PPB	11 Jan 2022 13:43:05	S	Std
0.2 - 1	50.2	987	-	0.20007471B	PPB	11 Jan 2022 13:45:22	S	Std
2.5 - 1	52.5	12180	-	2.50007471B	PPB	11 Jan 2022 13:47:39	S	Std
5.0 - 1	55	23688	-	5.00007471B	PPB	11 Jan 2022 13:49:56	S	Std
7.5 - 1	57.5	35911	-	7.50007471B	PPB	11 Jan 2022 13:54:40	S	Std
10.0 - 1	510	47850	-	10.00007471B	PPB	11 Jan 2022 13:59:30	S	Std
ICV40 - 1	ICV40	19657	4.1055	-7471B	PPB	11 Jan 2022 14:02:09	U	SMPL
ICB40 - 1	ICB40	-53	-0.0212	-7471B	PPB	11 Jan 2022 14:04:24	U	SMPL
CCV78 - 1	CCV78	23116	4.8298	-7471B	PPB	11 Jan 2022 14:06:40	U	SMPL
CCB78 - 1	CCB78	-52	-0.0210	-7471B	PPB	11 Jan 2022 14:08:55	U	SMPL
CRA - 1	CRA	959	0.1907	-7471B	PPB	11 Jan 2022 14:11:10	U	SMPL
HighStd - 1	HighStd	48257	10.0936	-7471B	PPB	11 Jan 2022 14:13:25	U	SMPL
ChkStd - 1	ChkStd	34040	7.1169	-7471B	PPB	11 Jan 2022 14:15:41	U	SMPL
PB141952BL - 1	PBS	-95	-0.0300	-7471B	PPB	11 Jan 2022 14:17:58	U	SMPL
PB141952BS - 1	LCSS	19822	4.1401	-7471B	PPB	11 Jan 2022 14:24:23	U	SMPL
N1054-01 - 1	20220014-CAPE-1-COMP	4038	0.8354	-7471B	PPB	11 Jan 2022 14:26:39	U	SMPL
N1054-02 - 1	20220015-CAPE-2-COMP	5623	1.1672	-7471B	PPB	11 Jan 2022 14:28:55	U	SMPL
N1054-03 - 1	20220016-CAPE-3-COMP	5969	1.2397	-7471B	PPB	11 Jan 2022 14:31:12	U	SMPL
N1054-04 - 1	20220017-CAPE-4-COMP	5272	1.0937	-7471B	PPB	11 Jan 2022 14:33:29	U	SMPL
N1054-05 - 1	20211303-CAPE-5-COMP	5381	1.1165	-7471B	PPB	11 Jan 2022 14:35:44	U	SMPL
CCV79 - 1	CCV79	23418	4.8930	-7471B	PPB	11 Jan 2022 14:38:00	U	SMPL
CCB79 - 1	CCB79	-50	-0.0206	-7471B	PPB	11 Jan 2022 14:40:16	U	SMPL
N1054-06 - 1	20220018-CAPE-6-COMP	6883	1.4310	-7471B	PPB	11 Jan 2022 14:42:31	U	SMPL
N1076-01 - 1	SB-01	10484	2.1850	-7471B	PPB	11 Jan 2022 14:56:45	U	SMPL
N1076-01DUP - 1	SB-01DUP	10018	2.0874	-7471B	PPB	11 Jan 2022 14:59:00	U	SMPL
N1076-01MS - 1	SB-01MS	29809	6.2311	-7471B	PPB	11 Jan 2022 15:01:15	U	SMPL
N1076-01MSD - 1	SB-01MSD	29917	6.2537	-7471B	PPB	11 Jan 2022 15:03:31	U	SMPL
N1076-04 - 1	SB-04	18323	3.8262	-7471B	PPB	11 Jan 2022 15:05:49	U	SMPL
N1076-07 - 1	SB-07	5704	1.1842	-7471B	PPB	11 Jan 2022 15:08:07	U	SMPL
N1076-13 - 1	SURFACE-01	8814	1.8353	-7471B	PPB	11 Jan 2022 15:10:24	U	SMPL
N1080-02 - 1	289	924	0.1834	-7471B	PPB	11 Jan 2022 15:12:40	U	SMPL
N1092-01 - 1	OK-03-011022	1533	0.3109	-7471B	PPB	11 Jan 2022 15:14:56	U	SMPL
CCV80 - 1	CCV80	23422	4.8938	-7471B	PPB	11 Jan 2022 15:17:11	U	SMPL
CCB80 - 1	CCB80	-40	-0.0185	-7471B	PPB	11 Jan 2022 15:19:27	U	SMPL
N1076-01LX5 - 1		2001	0.4089	-7471B	PPB	11 Jan 2022 15:21:43	U	SMPL
N1076-01A - 1		28984	6.0584	-7471B	PPB	11 Jan 2022 15:23:59	U	SMPL
CCV81 - 1	CCV81	22131	4.6235	-7471B	PPB	11 Jan 2022 15:26:14	U	SMPL
CCB81 - 1	CCB81	-28	-0.0159	-7471B	PPB	11 Jan 2022 15:28:30	U	SMPL