

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

1/11/2022 16:26

Reviewed by : 12 Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.913	0.0	0.108	
ICB1	-0.010	0.0	0.006	
CCV1	0.913	0.0	0.108	
CCB1	-0.009	0.0	0.006	
PB141895BL	-0.007	0.0	0.007	
PB141895BS	0.887	0.0	0.105	
N1057-01	-0.006	0.0	0.007	
N1057-01DUP	-0.003	0.0	0.007	
N1057-01MS	1.000	0.0	0.118	
N1057-01MSD	0.998	0.0	0.117	
N1058-01	-0.015	0.0	0.006	
N1058-02	-0.015	0.0	0.006	
CCV2	1.014	0.0	0.119	
CCB2	-0.007	0.0	0.007	

N 14
Mean 0.404
SD 0.4959
CV% 122.84

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 11 14:51:21 2022

Tue Jan 11 16:07:29 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0027	mg/l	1/11/2022 14:51:21	
0.05PPM	A	Phenolics-	P	0.0521	mg/l	1/11/2022 14:51:22	
0.1PPM	A	Phenolics-	P	0.1124	mg/l	1/11/2022 14:51:23	
0.25PPM	A	Phenolics-	P	0.2441	mg/l	1/11/2022 14:51:24	
0.50PPM	A	Phenolics-	P	0.4966	mg/l	1/11/2022 14:51:25	
1.0PPM	A	Phenolics-	P	0.9931	mg/l	1/11/2022 14:51:26	
2.0PPM	A	Phenolics-	P	2.0044	mg/l	1/11/2022 14:51:27	
ICV1	S	Phenolics-	P	0.913	mg/l	1/11/2022 15:17:59	
ICB1	S	Phenolics-	P	-0.0099	mg/l	1/11/2022 15:18:00	
CCV1	S	Phenolics-	P	0.9129	mg/l	1/11/2022 15:18:03	
CCB1	S	Phenolics-	P	-0.0095	mg/l	1/11/2022 15:18:05	
PB141895BL	S	Phenolics-	P	-0.0072	mg/l	1/11/2022 15:18:06	
PB141895BS	S	Phenolics-	P	0.8872	mg/l	1/11/2022 15:18:07	
N1057-01	S	Phenolics-	P	-0.0058	mg/l	1/11/2022 15:18:08	
N1057-01DUP	S	Phenolics-	P	-0.0034	mg/l	1/11/2022 15:18:09	
N1057-01MS	S	Phenolics-	P	1.0001	mg/l	1/11/2022 15:23:48	
N1057-01MSD	S	Phenolics-	P	0.9976	mg/l	1/11/2022 15:23:49	
N1058-01	S	Phenolics-	P	-0.0152	mg/l	1/11/2022 15:23:50	
N1058-02	S	Phenolics-	P	-0.0151	mg/l	1/11/2022 15:23:51	
CCV2	S	Phenolics-	P	1.014	mg/l	1/11/2022 16:07:27	
CCB2	S	Phenolics-	P	-0.0072	mg/l	1/11/2022 16:07:28	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

1/11/2022 14:52


Test Phenolics-

Accepted 1/11/2022 14:52

Factor 9.064
Bias 0.007

Coeff. of det. 0.999912

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.007	-0.0027	0.0000	-
2	Phenol-2	0.013	0.0521	0.0500	4.2
3	Phenol-2	0.020	0.1124	0.1000	12.4
4	Phenol-2	0.034	0.2441	0.2500	-2.4
5	Phenol-2	0.062	0.4966	0.5000	-0.7
6	Phenol-2	0.117	0.9931	1.0000	-0.7
7	Phenol-2	0.228	2.0044	2.0000	0.2

12
01/11/22