



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118164

Slope : 98.4

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/11/2022	17:00
2	CAL2	1	Water	NA	NA	20.6	7.00	01/11/2022	17:01
3	CAL3	1	Water	NA	NA	20.6	10.03	01/11/2022	17:04
4	ICV	1	Water	NA	NA	20.7	7.01	01/11/2022	17:05
5	CCV1	1	Water	NA	NA	20.6	2.01	01/11/2022	17:08
6	N1098-03	1	Water	NA	NA	20.1	7.43	01/11/2022	17:10
7	N1098-03DUP	1	Water	NA	NA	20.2	7.44	01/11/2022	17:11
8	N1102-01	1	Water	NA	NA	20.6	7.62	01/11/2022	17:13
9	N1102-02	1	Water	NA	NA	20.6	7.13	01/11/2022	17:15
10	CCV2	1	Water	NA	NA	20.7	12.03	01/11/2022	17:19

LB 118164

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1097 WorkList ID : 155919 Department : Wet-Chemistry Date : 01-11-2022 16:37:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/11/2022	Water	N1098-03	pH	Cool 4 deg C	PSEG03	I11	14-G-WATER	01/11/2022	9040C
01/11/2022	Water	N1102-01	pH	Cool 4 deg C	PSEG03	I11	31280	01/11/2022	9040C
01/11/2022	Water	N1102-02	pH	Cool 4 deg C	PSEG03	I11	31295	01/11/2022	9040C

Date/Time 01.11.22 16:40
 Raw Sample Received by: JN (WC)
 Raw Sample Relinquished by: JN (WC)

Date/Time 01.11.22 18:20
 Raw Sample Received by: JN (WC)
 Raw Sample Relinquished by: JN (WC)