

LB118170


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35923



A= 0.0000e+000

B= 2.8357e-004

C= 1.3427e-002

Rho= 0.9991301

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	59	0.000	59					1
0.2	0.200	0.216	0.016	714	0.0 %	714					8
2.5	2.500	2.649	0.149	9294	0.0 %	9294					6
5.0	5.000	4.822	-0.178	16958	0.0 %	16958					-4
7.5	7.500	7.283	-0.217	25636	0.0 %	25636					-3
10.0	10.000	10.200	0.200	35923	0.0 %	35923					2

LB118170 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jan 2022 09:16:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	59	-	0.0000	7470A	PPB	12 Jan 2022 09:26:21	S	Std
0.2 - 1	50.2	714	-	0.2000	7470A	PPB	12 Jan 2022 09:32:32	S	Std
2.5 - 1	52.5	9294	-	2.5000	7470A	PPB	12 Jan 2022 09:34:49	S	Std
5.0 - 1	55	16958	-	5.0000	7470A	PPB	12 Jan 2022 09:37:06	S	Std
7.5 - 1	57.5	25636	-	7.5000	7470A	PPB	12 Jan 2022 09:42:04	S	Std
10.0 - 1	58	35923	-	10.0000	7470A	PPB	12 Jan 2022 09:46:56	S	Std
ICV41 - 1	ICV41	13792	3.9244	-	7470A	PPB	12 Jan 2022 09:49:38	U	SMPL
ICB41 - 1	ICB41	-40	0.0021	-	7470A	PPB	12 Jan 2022 09:51:53	U	SMPL
CCV82 - 1	CCV82	15824	4.5006	-	7470A	PPB	12 Jan 2022 09:54:08	U	SMPL
CCB82 - 1	CCB82	-37	0.0029	-	7470A	PPB	12 Jan 2022 09:56:23	U	SMPL
CRA - 1	CRA	656	0.1994	-	7470A	PPB	12 Jan 2022 09:58:38	U	SMPL
HighStd - 1	HighStd	36031	10.2306	-	7470A	PPB	12 Jan 2022 10:00:53	U	SMPL
ChkStd - 1	ChkStd	23798	6.7617	-	7470A	PPB	12 Jan 2022 10:03:09	U	SMPL
PB141972BL - 1	PBW	-51	-0.0010	-	7470A	PPB	12 Jan 2022 10:05:25	U	SMPL
PB141972BS - 1	LCSW	13604	3.8711	-	7470A	PPB	12 Jan 2022 10:11:48	U	SMPL
N1087-01 - 1	FMI-114	1040	0.3083	-	7470A	PPB	12 Jan 2022 10:14:04	U	SMPL
N1087-01DUP - 1	FMI-114DUP	977	0.2905	-	7470A	PPB	12 Jan 2022 10:16:20	U	SMPL
N1087-01MS - 1	FMI-114MS	14634	4.1631	-	7470A	PPB	12 Jan 2022 10:21:07	U	SMPL
N1087-01MSD - 1	FMI-114MSD	14878	4.2323	-	7470A	PPB	12 Jan 2022 10:23:24	U	SMPL
CCV83 - 1	CCV83	16432	4.6730	-	7470A	PPB	12 Jan 2022 10:25:39	U	SMPL
CCB83 - 1	CCB83	-64	-0.0047	-	7470A	PPB	12 Jan 2022 10:27:55	U	SMPL
N1087-01LX5 - 1		185	0.0659	-	7470A	PPB	12 Jan 2022 10:30:10	U	SMPL
N1087-01A - 1		14960	4.2556	-	7470A	PPB	12 Jan 2022 10:32:25	U	SMPL
CCV84 - 1	CCV84	18057	5.1338	-	7470A	PPB	12 Jan 2022 10:34:41	U	SMPL
CCB84 - 1	CCB84	-56	-0.0025	-	7470A	PPB	12 Jan 2022 10:36:56	U	SMPL