

LB118172


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

47495



A= 0.0000e+000

B= 2.0963e-004

C= 1.4320e-002

Rho= 0.9987188

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.020	0.020	29	0.000	29					
0.2	0.200	0.213	0.013	949	0.0 %	949					
2.5	2.500	2.664	0.164	12638	0.0 %	12638					
5.0	5.000	4.627	-0.373	22005	0.0 %	22005					
7.5	7.500	7.705	0.205	36687	0.0 %	36687					
10.0	10.000	9.971	-0.029	47495	0.0 %	47495					

6
7
-7
3
0

LB118172 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Jan 2022 11:06:48

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	29	-	0.00007471B	PPB	12 Jan 2022 11:07:42	S	Std
0.2 - 1	50.2	949	-	0.20007471B	PPB	12 Jan 2022 11:09:59	S	Std
2.5 - 1	52.5	12638	-	2.50007471B	PPB	12 Jan 2022 11:12:16	S	Std
5.0 - 1	55	22005	-	5.00007471B	PPB	12 Jan 2022 11:14:33	S	Std
7.5 - 1	57.5	36687	-	7.50007471B	PPB	12 Jan 2022 11:19:20	S	Std
10.0 - 1	511	47495	-	10.00007471B	PPB	12 Jan 2022 11:24:09	S	Std
ICV42 - 1	ICV42	20317	4.2733	-7471B	PPB	12 Jan 2022 11:26:53	U	SMPL
ICB42 - 1	ICB42	-47	0.0045	-7471B	PPB	12 Jan 2022 11:29:08	U	SMPL
CCV85 - 1	CCV85	23184	4.8743	-7471B	PPB	12 Jan 2022 11:31:23	U	SMPL
CCB85 - 1	CCB85	-37	0.0066	-7471B	PPB	12 Jan 2022 11:33:38	U	SMPL
CRA - 1	CRA	907	0.2045	-7471B	PPB	12 Jan 2022 11:35:54	U	SMPL
HighStd - 1	HighStd	47202	9.9092	-7471B	PPB	12 Jan 2022 11:38:09	U	SMPL
ChkStd - 1	ChkStd	34532	7.2532	-7471B	PPB	12 Jan 2022 11:40:25	U	SMPL
PB141973BL - 1	PBS	-96	-0.0058	-7471B	PPB	12 Jan 2022 11:42:42	U	SMPL
N1097-01 - 1	TP-1S	363595	HIGH	-7471B	PPB	12 Jan 2022 11:47:15	X	
N1097-04 - 1	TP-1D	210828	44.2098	-7471B	PPB	12 Jan 2022 11:49:06	U	SMPL
N1097-07 - 1	TP-2S	58948	12.3715	-7471B	PPB	12 Jan 2022 11:51:25	U	SMPL
N1097-10 - 1	TP-2D	5341	1.1339	-7471B	PPB	12 Jan 2022 11:53:45	U	SMPL
N1097-13 - 1	TP-3S	340128	HIGH	-7471B	PPB	12 Jan 2022 11:56:02	X	
CCV86 - 1	CCV86	23013	4.8385	-7471B	PPB	12 Jan 2022 11:57:52	U	SMPL
CCB86 - 1	CCB86	-120	-0.0108	-7471B	PPB	12 Jan 2022 12:00:10	U	SMPL
N1097-16 - 1	TP-3D	46340	9.7285	-7471B	PPB	12 Jan 2022 12:02:26	U	SMPL
N1099-01 - 1	VACLOT-BACKFILL-TILCON-01	-35	0.0070	-7471B	PPB	12 Jan 2022 12:04:41	U	SMPL
N1099-02 - 1	VACLOT-BACKFILL-TILCON-02	44	0.0235	-7471B	PPB	12 Jan 2022 12:06:58	U	SMPL
N1099-03 - 1	VACLOT-BACKFILL-TILCON-03	57	0.0263	-7471B	PPB	12 Jan 2022 12:09:14	U	SMPL
N1099-04 - 1	VACLOT-BACKFILL-TILCON-04	66	0.0282	-7471B	PPB	12 Jan 2022 12:11:29	U	SMPL
N1099-05 - 1	VACLOT-BACKFILL-TILCON-05	88	0.0328	-7471B	PPB	12 Jan 2022 12:13:46	U	SMPL
N1099-06 - 1	VACLOT-BACKFILL-TILCON-06	93	0.0338	-7471B	PPB	12 Jan 2022 12:16:03	U	SMPL
N1099-07 - 1	VACLOT-BACKFILL-TILCON-07	82	0.0315	-7471B	PPB	12 Jan 2022 12:18:19	U	SMPL
N1099-08 - 1	VACLOT-BACKFILL-TILCON-08	48	0.0244	-7471B	PPB	12 Jan 2022 12:20:36	U	SMPL
N1099-09 - 1	VACLOT-BACKFILL-TILCON-09	59	0.0267	-7471B	PPB	12 Jan 2022 12:22:52	U	SMPL
CCV87 - 1	CCV87	22898	4.8144	-7471B	PPB	12 Jan 2022 12:25:07	U	SMPL
CCB87 - 1	CCB87	-32	0.0076	-7471B	PPB	12 Jan 2022 12:27:23	U	SMPL
N1099-10 - 1	VACLOT-BACKFILL-TILCON-10	77	0.0305	-7471B	PPB	12 Jan 2022 12:29:39	U	SMPL
N1103-01 - 1	CL-01-011122	4079	0.8694	-7471B	PPB	12 Jan 2022 12:31:54	U	SMPL
N1103-01DUP - 1	CL-01-011122DUP	4001	0.8530	-7471B	PPB	12 Jan 2022 12:34:10	U	SMPL
N1103-01MS - 1	CL-01-011122MS	24630	5.1775	-7471B	PPB	12 Jan 2022 12:36:25	U	SMPL
N1103-01MSD - 1	CL-01-011122MSD	23714	4.9854	-7471B	PPB	12 Jan 2022 12:40:37	U	SMPL
N1107-01 - 1	PT-03-011122	3802	0.8113	-7471B	PPB	12 Jan 2022 12:42:53	U	SMPL
N1107-03 - 1	PT-04-011122	1589	0.3474	-7471B	PPB	12 Jan 2022 12:45:09	U	SMPL
N1103-01LX5 - 1		752	0.1720	-7471B	PPB	12 Jan 2022 12:47:26	U	SMPL
N1103-01A - 1		23756	4.9942	-7471B	PPB	12 Jan 2022 12:49:43	U	SMPL
PB141973BS - 1	LCSS	18739	3.9425	-7471B	PPB	12 Jan 2022 12:51:59	U	SMPL
CCV88 - 1	CCV88	22907	4.8163	-7471B	PPB	12 Jan 2022 12:54:15	U	SMPL
CCB88 - 1	CCB88	-35	0.0070	-7471B	PPB	12 Jan 2022 12:56:30	U	SMPL
N1097-01DLX10 - 1		49166	10.3209	-7471B	PPB	12 Jan 2022 12:58:46	U	SMPL
N1097-04DLX10 - 1		21982	4.6224	-7471B	PPB	12 Jan 2022 13:01:01	U	SMPL
N1097-07DLX2 - 1		29223	6.1403	-7471B	PPB	12 Jan 2022 13:03:20	U	SMPL
N1097-13DLX10 - 1		52701	11.0619	-7471B	PPB	12 Jan 2022 13:05:35	U	SMPL
N1097-01DL2X25 - 1		21804	4.5851	-7471B	PPB	12 Jan 2022 13:07:51	U	SMPL
N1097-13DL2X25 - 1		19692	4.1423	-7471B	PPB	12 Jan 2022 13:10:09	U	SMPL
CCV89 - 1	CCV89	23044	4.8450	-7471B	PPB	12 Jan 2022 13:12:25	U	SMPL
CCB89 - 1	CCB89	-48	0.0043	-7471B	PPB	12 Jan 2022 13:14:42	U	SMPL