



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/31/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 01/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:18
Out Date: 01/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118184

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1108-01	BG6T6	1	1.11	8.87	9.98	8.76	86.2	
N1108-02	BGXR5	2	1.14	8.63	9.77	8.68	87.4	
N1108-03	BGXX7	3	1.13	8.69	9.82	5.97	55.7	
N1108-04	BGXZ5	4	1.19	8.66	9.85	8.97	89.8	
N1108-05	BGY01	5	1.15	8.82	9.97	7.59	73.0	
N1108-06	BGY24	6	1.13	8.53	9.66	6.86	67.2	
N1108-07	BGY25	7	1.13	8.54	9.67	6.34	61.0	
N1108-08	BGY27	8	1.15	8.81	9.96	7.06	67.1	
N1108-09	BGY32	9	1.19	8.43	9.62	7.62	76.3	
N1108-10	BGY36	10	1.15	8.56	9.71	6.26	59.7	
N1108-11	BGY41	11	1.14	8.82	9.96	5.22	46.3	
N1108-12	BGY42	12	1.19	8.65	9.84	5.56	50.5	
N1108-13	BGY51	13	1.19	8.55	9.74	7.93	78.8	
N1108-14	BGY54	14	1.19	8.64	9.83	8.18	80.9	
N1108-15	BGY61	15	1.15	8.81	9.96	8.21	80.1	
N1108-16	BGY62	16	1.16	8.34	9.5	7.54	76.5	
N1108-17	BGY63	17	1.16	8.81	9.97	7.66	73.8	
N1108-18	BGY66	18	1.14	8.81	9.95	7.7	74.5	
N1108-19	BGY80	19	1.13	8.44	9.57	8.55	87.9	
N1108-20	BGYQ4	20	1.14	8.80	9.94	8.95	88.8	
N1108-21	BGYQ4D	21	1.14	8.80	9.94	8.95	88.8	
N1108-22	BGYQ4S	22	1.14	8.80	9.94	8.95	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$