



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/13/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 01/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 01/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118185

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1109-01	BGY93	1	1.14	8.80	9.94	8.81	87.2	
N1109-02	BGY99	2	1.16	8.62	9.78	8.54	85.6	
N1109-03	BGYC0	3	1.14	8.82	9.96	9.36	93.2	
N1109-04	BGYC4	4	1.16	8.62	9.78	7.87	77.8	
N1109-05	BGYC7	5	1.19	8.46	9.65	8.74	89.2	
N1109-06	BGYD3	6	1.14	8.41	9.55	6.85	67.9	
N1109-07	BGYD4	7	1.12	8.45	9.57	6.29	61.2	
N1109-08	BGYD8	8	1.11	8.66	9.77	8.45	84.8	
N1109-09	BGYE4	9	1.19	8.71	9.9	8.19	80.4	
N1109-10	BGYE5	10	1.16	8.53	9.69	8.78	89.3	
N1109-11	BGYF0	11	1.19	8.73	9.92	8.71	86.1	
N1109-12	BGYF1	12	1.11	8.41	9.52	8.71	90.4	
N1109-13	BGYK8	13	1.17	8.71	9.88	8.99	89.8	
N1109-14	BGYK9	14	1.15	8.83	9.98	8.67	85.2	
N1109-15	BGYM1	15	1.16	8.71	9.87	8.44	83.6	
N1109-16	BGYM2	16	1.16	8.66	9.82	8.45	84.2	
N1109-17	BGYM4	17	1.16	8.81	9.97	9.34	92.8	
N1109-18	BGYM8	18	1.12	8.41	9.53	9.05	94.3	
N1109-19	BGYN2	19	1.15	8.80	9.95	9.54	95.3	
N1109-20	BGYQ5	20	1.16	8.59	9.75	9.37	95.6	
N1109-21	BGYQ5D	21	1.16	8.59	9.75	9.37	95.6	
N1109-22	BGYQ5S	22	1.16	8.59	9.75	9.37	95.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$