



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: Iwona

Parameter: Color

Supervisor Review By: apatel

Run Number: LB118186

pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP94763
Calibration std, 5 color unit	WP94764
Calibration std, 10 color unit	WP94765
Calibration std, 15 color unit	WP94766
Calibration std, 20 color unit	WP94767
Calibration std, 25 color unit	WP94768
Calibration std, 30 color unit	WP94769
Calibration std, 40 color unit	WP94770
Calibration std, 50 color unit	WP94771
Calibration std, 60 color unit	WP94772
Calibration std, 70 color unit	WP94773

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB118186BL	LB118186BL	1	5.10	N/A	0.00	01/12/2022	11:20
2	N1104-07	7CS	10	8.04	N/A	50.00	01/12/2022	11:25
3	N1104-07DUP	7CSDUP	10	7.98	N/A	50.00	01/12/2022	11:30

WORKLIST(Hardcopy Internal Chain)

LB 118186

WorkList Name : COLOR-011222 WorkList ID : 155946 Department : Wet-Chemistry Date : 01-12-2022 09:37:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/13/2022	Water	N1104-07	Color	Cool 4 deg C	M&MM01	H11	7CS	01/11/2022	SM2120 B

Date/Time 01/12/22 10:00
 Raw Sample Received by: RLC
 Raw Sample Relinquished by: KS (WC)

Date/Time 01/12/22 12:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RLC