


LB118188

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

45470



Linear

A= 0.0000e+000

B= 2.2893e-004 *slope*C= -3.5979e-001 *y-intercept*

Rho= 0.9999073

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>@/oD</i>
0.00	0.000	0.021	0.021	1663	0.000	1663					=
0.05	0.050					205	0				-16
0.20	0.200	0.168	-0.032	2304	0.0 %	2304					1
2.50	2.500	2.528	0.028	12616	0.0 %	12616					1
5.00	5.000	5.031	0.031	23546	0.0 %	23546					1
7.50	7.500	7.403	-0.097	33908	0.0 %	33908					1
10.0	10.000	10.050	0.050	45470	0.0 %	45470					1

LB118188
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	1663	-		0 SFAM01.1	PPB	1/12/2022 14:51	Std
	0.2 S01	2304	-		0.2 SFAM01.1	PPB	1/12/2022 14:53	Std
	2.5 S02	12616	-		2.5 SFAM01.1	PPB	1/12/2022 14:56	Std
	5 S03	23546	-		5 SFAM01.1	PPB	1/12/2022 14:58	Std
	7.5 S04	33908	-		7.5 SFAM01.1	PPB	1/12/2022 15:00	Std
	10 S05	45470	-		10 SFAM01.1	PPB	1/12/2022 15:03	Std
ICV007	ICV007	18468	3.8681 -		SFAM01.1	PPB	1/12/2022 15:05	SMPL
ICB007	ICB007	1615	0.0099 -		SFAM01.1	PPB	1/12/2022 15:08	SMPL
CCV090	CCV090	22940	4.8919 -		SFAM01.1	PPB	1/12/2022 15:10	SMPL
CCB090	CCB090	1611	0.009 -		SFAM01.1	PPB	1/12/2022 15:12	SMPL
PB141993BL	PBW993	1578	0.0015 -		SFAM01.1	PPB	1/12/2022 15:17	SMPL
N1028-01	EW306	59	-0.3463 -		SFAM01.1	PPB	1/12/2022 15:19	SMPL
N1028-02	EW309	43	-0.3499 -		SFAM01.1	PPB	1/12/2022 15:21	SMPL
N1028-03	EW311	52	-0.3479 -		SFAM01.1	PPB	1/12/2022 15:24	SMPL
N1028-04	EW313	23	-0.3545 -		SFAM01.1	PPB	1/12/2022 15:26	SMPL
N1028-05	EW314	80	-0.3415 -		SFAM01.1	PPB	1/12/2022 15:28	SMPL
N1028-06	EW320	65	-0.3449 -		SFAM01.1	PPB	1/12/2022 15:30	SMPL
N1028-07	EW2W5	50	-0.3483 -		SFAM01.1	PPB	1/12/2022 15:33	SMPL
N1028-08	EW304	50	-0.3483 -		SFAM01.1	PPB	1/12/2022 15:35	SMPL
N1028-09	EW304D	47	-0.349 -		SFAM01.1	PPB	1/12/2022 15:37	SMPL
N1028-10	EW304S	5806	0.9694 -		SFAM01.1	PPB	1/12/2022 15:44	SMPL
N1028-11	EW305	55	-0.3472 -		SFAM01.1	PPB	1/12/2022 15:47	SMPL
N1028-12	EW307	68	-0.3442 -		SFAM01.1	PPB	1/12/2022 15:49	SMPL
N1028-13	EW308	80	-0.3415 -		SFAM01.1	PPB	1/12/2022 15:51	SMPL
N1028-14	EW312	56	-0.347 -		SFAM01.1	PPB	1/12/2022 15:53	SMPL
PB141994BL	PBW994	1542	-0.0068 -		SFAM01.1	PPB	1/12/2022 15:56	SMPL
N1075-01	EW2W6	49	-0.3486 -		SFAM01.1	PPB	1/12/2022 15:58	SMPL
N1075-02	EW310	49	-0.3486 -		SFAM01.1	PPB	1/12/2022 16:00	SMPL
N1075-03	EW315	56	-0.347 -		SFAM01.1	PPB	1/12/2022 16:03	SMPL
N1075-04	EW315D	58	-0.3465 -		SFAM01.1	PPB	1/12/2022 16:05	SMPL
CCV091	CCV091	22385	4.7648 -		SFAM01.1	PPB	1/12/2022 16:07	SMPL
CCB091	CCB091	1518	-0.0123 -		SFAM01.1	PPB	1/12/2022 16:09	SMPL
N1075-05	EW315S	5973	1.0076 -		SFAM01.1	PPB	1/12/2022 16:14	SMPL
N1075-06	EW318	236	-0.3058 -		SFAM01.1	PPB	1/12/2022 16:16	SMPL
N1075-07	EW319	120	-0.3323 -		SFAM01.1	PPB	1/12/2022 16:19	SMPL
N1075-08	EW321	107	-0.3353 -		SFAM01.1	PPB	1/12/2022 16:21	SMPL
N1075-09	EW316	165	-0.322 -		SFAM01.1	PPB	1/12/2022 16:23	SMPL
N1075-10	EW317	38	-0.3511 -		SFAM01.1	PPB	1/12/2022 16:25	SMPL
CCV092	CCV092	21598	4.5846 -		SFAM01.1	PPB	1/12/2022 16:28	SMPL
CCB092	CCB092	1489	-0.0189 -		SFAM01.1	PPB	1/12/2022 16:30	SMPL