


LB118199

7470A

INSTRUMENT ID: CV4

μ Abs.:

40797



Linear

A= 0.0000e+000

B= 2.4655e-004

C= -4.6086e-002

Rho= 0.9998710

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0)
0.0	0.000	-0.049	-0.049	-11	0.000	-11					-
0.2	0.200	0.187	-0.013	947	0.0 %	947					7
2.5	2.500	2.564	0.064	10585	0.0 %	10585					3
5.0	5.000	5.077	0.077	20778	0.0 %	20778					2
7.5	7.500	7.408	-0.092	30234	0.0 %	30234					-1
10.0	10.000	10.013	0.013	40797	0.0 %	40797					0

LB118199 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jan 2022 10:44:20

Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.00007470A	PPB	13 Jan 2022 10:54:56	S	Std
0.2 - 1	50.2	947	-	0.20007470A	PPB	13 Jan 2022 10:57:13	S	Std
2.5 - 1	52.5	10585	-	2.50007470A	PPB	13 Jan 2022 10:59:30	S	Std
5.0 - 1	55	20778	-	5.00007470A	PPB	13 Jan 2022 11:01:47	S	Std
7.5 - 1	57.5	30234	-	7.50007470A	PPB	13 Jan 2022 11:04:03	S	Std
10.0 - 1	50	40797	-	10.00007470A	PPB	13 Jan 2022 11:10:23	S	Std
ICV43 - 1	ICV43	15226	3.7080	-7470A	PPB	13 Jan 2022 11:15:33	U	SMPL
ICB43 - 1	ICB43	81	-0.0261	-7470A	PPB	13 Jan 2022 11:17:48	U	SMPL
CCV93 - 1	CCV93	19301	4.7127	-7470A	PPB	13 Jan 2022 11:20:03	U	SMPL
CCB93 - 1	CCB93	65	-0.0301	-7470A	PPB	13 Jan 2022 11:22:19	U	SMPL
CRA - 1	CRA	945	0.1869	-7470A	PPB	13 Jan 2022 11:24:33	U	SMPL
HighStd - 1	HighStd	40923	10.0437	-7470A	PPB	13 Jan 2022 11:26:49	U	SMPL
ChkStd - 1	ChkStd	27058	6.6252	-7470A	PPB	13 Jan 2022 11:29:04	U	SMPL
PB141991BL - 1	PBW	228	0.0101	-7470A	PPB	13 Jan 2022 11:31:21	U	SMPL
PB141991BS - 1	LCSW	15053	3.6653	-7470A	PPB	13 Jan 2022 11:36:08	U	SMPL
N1097-02 - 1	TP-1S	347	0.0395	-7470A	PPB	13 Jan 2022 11:38:24	U	SMPL
N1097-02DUP - 1	TP-1SDUP	354	0.0412	-7470A	PPB	13 Jan 2022 11:40:40	U	SMPL
N1097-02MS - 1	TP-1SMS	17451	4.2565	-7470A	PPB	13 Jan 2022 11:42:57	U	SMPL
N1097-02MSD - 1	TP-1SMSD	17765	4.3340	-7470A	PPB	13 Jan 2022 11:45:14	U	SMPL
N1097-05 - 1	TP-1D	330	0.0353	-7470A	PPB	13 Jan 2022 11:47:29	U	SMPL
CCV94 - 1	CCV94	19669	4.8034	-7470A	PPB	13 Jan 2022 11:49:44	U	SMPL
CCB94 - 1	CCB94	65	-0.0301	-7470A	PPB	13 Jan 2022 11:51:59	U	SMPL
N1097-08 - 1	TP-2S	589	0.0991	-7470A	PPB	13 Jan 2022 11:54:15	U	SMPL
N1097-11 - 1	TP-2D	661	0.1169	-7470A	PPB	13 Jan 2022 11:59:52	U	SMPL
N1097-14 - 1	TP-3S	3158	0.7325	-7470A	PPB	13 Jan 2022 12:02:07	U	SMPL
N1097-17 - 1	TP-3D	900	0.1758	-7470A	PPB	13 Jan 2022 12:06:58	U	SMPL
PB141992BL - 1	PBW	-11	-0.0488	-7470A	PPB	13 Jan 2022 12:09:13	U	SMPL
PB141992BS - 1	LCSW	14154	3.4437	-7470A	PPB	13 Jan 2022 12:21:10	U	SMPL
N1099-11 - 1	VACLOT-BACKFILL-TILCON-01	230	0.0106	-7470A	PPB	13 Jan 2022 12:23:26	U	SMPL
N1099-12 - 1	VACLOT-BACKFILL-TILCON-02	527	0.0838	-7470A	PPB	13 Jan 2022 12:25:43	U	SMPL
N1099-13 - 1	VACLOT-BACKFILL-TILCON-03	380	0.0476	-7470A	PPB	13 Jan 2022 12:28:00	U	SMPL
N1099-14 - 1	VACLOT-BACKFILL-TILCON-04	511	0.0799	-7470A	PPB	13 Jan 2022 12:30:16	U	SMPL
CCV95 - 1	CCV95	19255	4.7013	-7470A	PPB	13 Jan 2022 12:32:32	U	SMPL
CCB95 - 1	CCB95	-3	-0.0468	-7470A	PPB	13 Jan 2022 12:34:48	U	SMPL
N1099-15 - 1	VACLOT-BACKFILL-TILCON-05	388	0.0496	-7470A	PPB	13 Jan 2022 12:37:04	U	SMPL
N1099-16 - 1	VACLOT-BACKFILL-TILCON-06	406	0.0540	-7470A	PPB	13 Jan 2022 12:39:19	U	SMPL
N1099-17 - 1	VACLOT-BACKFILL-TILCON-07	374	0.0461	-7470A	PPB	13 Jan 2022 12:41:35	U	SMPL
N1099-18 - 1	VACLOT-BACKFILL-TILCON-08	389	0.0498	-7470A	PPB	13 Jan 2022 12:43:50	U	SMPL
N1099-19 - 1	VACLOT-BACKFILL-TILCON-09	394	0.0511	-7470A	PPB	13 Jan 2022 12:46:06	U	SMPL
N1099-20 - 1	VACLOT-BACKFILL-TILCON-10	428	0.0594	-7470A	PPB	13 Jan 2022 12:48:23	U	SMPL
N1099-20DUP - 1	VACLOT-BACKFILL-TILCON-10DUP	428	0.0594	-7470A	PPB	13 Jan 2022 12:50:39	U	SMPL
N1099-20MS - 1	VACLOT-BACKFILL-TILCON-10MS	16746	4.0827	-7470A	PPB	13 Jan 2022 12:52:56	U	SMPL
N1099-20MSD - 1	VACLOT-BACKFILL-TILCON-10MSD	16896	4.1197	-7470A	PPB	13 Jan 2022 12:55:13	U	SMPL
CCV96 - 1	CCV96	21792	5.3268	-7470A	PPB	13 Jan 2022 12:57:28	U	SMPL
CCB96 - 1	CCB96	85	-0.0251	-7470A	PPB	13 Jan 2022 13:02:23	U	SMPL
N1097-02LX5 - 1		64	-0.0303	-7470A	PPB	13 Jan 2022 13:04:39	U	SMPL
N1097-02A - 1		15326	3.7326	-7470A	PPB	13 Jan 2022 13:06:54	U	SMPL
N1099-20LX5 - 1		44	-0.0352	-7470A	PPB	13 Jan 2022 13:09:10	U	SMPL
N1099-20A - 1		14983	3.6480	-7470A	PPB	13 Jan 2022 13:11:26	U	SMPL
PB141974TB - 1	PB141974TB	327	0.0345	-7470A	PPB	13 Jan 2022 13:13:41	U	SMPL
PB141969TB - 1	PB141969TB	385	0.0488	-7470A	PPB	13 Jan 2022 13:15:57	U	SMPL
CCV97 - 1	CCV97	18724	4.5704	-7470A	PPB	13 Jan 2022 13:18:13	U	SMPL
CCB97 - 1	CCB97	66	-0.0298	-7470A	PPB	13 Jan 2022 13:20:29	U	SMPL

PB141992TB NB
11/26/22