



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118203

Slope : 98.4

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/13/2022	12:30
2	CAL2	1	Water	NA	NA	20.3	7.00	01/13/2022	12:31
3	CAL3	1	Water	NA	NA	20.3	10.03	01/13/2022	12:33
4	ICV	1	Water	NA	NA	20.4	6.99	01/13/2022	12:35
5	CCV1	1	Water	NA	NA	20.4	2.01	01/13/2022	12:40
6	N1128-01	1	Water	NA	NA	26.6	5.28	01/13/2022	12:43
7	N1128-01DUP	1	Water	NA	NA	26.7	5.29	01/13/2022	12:44
8	CCV2	1	Water	NA	NA	20.4	12.03	01/13/2022	12:45

WORKLIST(Hardcopy Internal Chain)

W118203

WorkList Name : PH N1128

WorkList ID : 155993

Department : Wet-Chemistry

Date : 01-13-2022 12:22:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2022	Water	N1128-01	pH	Cool 4 deg C	STER05	H11	BOTTLE-1	01/10/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01-13-22 12:25
70626
70626

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01-13-22 13:30
70626
70626