

L3118215


7471B

INSTRUMENT ID: EV1

Linear

μ Abs.:

52212



A= 0.0000e+000

B= 1.9088e-004

C= 1.5369e-003

Rho= 0.9999447

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.012	0.012	55	0.000	55					9
0.2	0.200	0.219	0.019	1141	0.0 %	1141					-2
2.5	2.500	2.438	-0.062	12764	0.0 %	12764					0
5.0	5.000	5.003	0.003	26202	0.0 %	26202					1
7.5	7.500	7.560	0.060	39599	0.0 %	39599					0
10.0	10.000	9.968	-0.032	52212	0.0 %	52212					

LB118215 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jan 2022 15:43:50

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	55	-	0.000	7471B	PPB	13 Jan 2022 15:44:44	S	Std
0.2 - 1	SO-2	1141	-	0.200	7471B	PPB	13 Jan 2022 15:47:01	S	Std
2.5 - 1	SO-5	12764	-	2.500	7471B	PPB	13 Jan 2022 15:49:18	S	Std
5.0 - 1	SO	26202	-	5.000	7471B	PPB	13 Jan 2022 15:51:35	S	Std
7.5 - 1	SO-5	39599	-	7.500	7471B	PPB	13 Jan 2022 15:53:51	S	Std
10.0 - 1	SO	52212	-	10.000	7471B	PPB	13 Jan 2022 15:58:48	S	Std
ICV44 - 1	ICV44	22030	4.2066	-	7471B	PPB	13 Jan 2022 16:06:34	U	SMPL
ICB44 - 1	ICB44	-52	-0.0084	-	7471B	PPB	13 Jan 2022 16:08:49	U	SMPL
CCV01 - 1	CCV01	24861	4.7470	-	7471B	PPB	13 Jan 2022 16:11:05	U	SMPL
CCB01 - 1	CCB01	-63	-0.0105	-	7471B	PPB	13 Jan 2022 16:13:20	U	SMPL
CRA - 1	CRA	972	0.1871	-	7471B	PPB	13 Jan 2022 16:15:36	U	SMPL
HighStd - 1	HighStd	53396	10.1937	-	7471B	PPB	13 Jan 2022 16:17:51	U	SMPL
ChkStd - 1	ChkStd	36312	6.9327	-	7471B	PPB	13 Jan 2022 16:20:07	U	SMPL
PB142019BL - 1	PBS	-46	-0.0072	-	7471B	PPB	13 Jan 2022 16:22:24	U	SMPL
PB142019BS - 1	LCSS	19782	3.7775	-	7471B	PPB	13 Jan 2022 16:24:41	U	SMPL
N1112-01 - 1	TP-4S	73413	14.0145	-	7471B	PPB	13 Jan 2022 16:26:57	U	SMPL
N1112-04 - 1	TP-4D	7514	1.4358	-	7471B	PPB	13 Jan 2022 16:29:14	U	SMPL
N1118-01 - 1	MR-SOILPILE	1093	0.2102	-	7471B	PPB	13 Jan 2022 16:31:34	U	SMPL
N1118-01DUP - 1	MR-SOILPILE	1058	0.2035	-	7471B	PPB	13 Jan 2022 16:33:50	U	SMPL
N1118-01MS - 1	MR-SOILPILEMS	19238	3.6737	-	7471B	PPB	13 Jan 2022 16:36:06	U	SMPL
CCV02 - 1	CCV02	25150	4.8021	-	7471B	PPB	13 Jan 2022 16:38:21	U	SMPL
CCB02 - 1	CCB02	-63	-0.0105	-	7471B	PPB	13 Jan 2022 16:40:36	U	SMPL
N1118-01MSD - 1	MR-SOILPILEMSD	18535	3.5395	-	7471B	PPB	13 Jan 2022 16:42:53	U	SMPL
N1119-01 - 1	NORTH-EAST-CORNER-EXALVATION-NORTH-SIDEWALL	11807	2.2552	-	7471B	PPB	13 Jan 2022 16:45:08	U	SMPL
N1119-02 - 1	NORTH-EAST-CORNER-EXALVATION-SOUTH-SIDEWALL	134	0.0271	-	7471B	PPB	13 Jan 2022 16:47:25	U	SMPL
N1119-03 - 1	NORTH-EAST-CORNER-EXALVATION-WEST-SIDEWALL	1713	0.3285	-	7471B	PPB	13 Jan 2022 16:49:41	U	SMPL
N1119-04 - 1	NORTH-EAST-CORNER-EXALVATION-EAST-SIDEWALL	2164	0.4146	-	7471B	PPB	13 Jan 2022 16:51:56	U	SMPL
N1119-05 - 1	EAST-AREA-BOTTOM	676	0.1306	-	7471B	PPB	13 Jan 2022 16:54:12	U	SMPL
N1119-06 - 1	EAST-AREA-EAST-SIDE-WALL	958	0.1844	-	7471B	PPB	13 Jan 2022 16:56:29	U	SMPL
N1120-01 - 1	EAST-BOILER-PIT-BOTTON	1114	0.2142	-	7471B	PPB	13 Jan 2022 16:58:46	U	SMPL
N1121-01 - 1	WASTE-CLASS-LAUNDRY-PIT	340296	HIGH	-	7471B	PPB	13 Jan 2022 17:01:02	X	
CCV03 - 1	CCV03	25039	4.7809	-	7471B	PPB	13 Jan 2022 17:02:53	U	SMPL
CCB03 - 1	CCB03	-127	-0.0227	-	7471B	PPB	13 Jan 2022 17:05:12	U	SMPL
N1118-01LX5 - 1		215	0.0426	-	7471B	PPB	13 Jan 2022 17:07:29	U	SMPL
N1118-01A - 1		20692	3.9512	-	7471B	PPB	13 Jan 2022 17:09:45	U	SMPL
N1112-01DLX5 - 1		15463	2.9531	-	7471B	PPB	13 Jan 2022 17:12:00	U	SMPL
N1121-01DLX10 - 1		58673	11.2009	-	7471B	PPB	13 Jan 2022 17:14:16	U	SMPL
N1121-01DLX25 - 1		23454	4.4784	-	7471B	PPB	13 Jan 2022 17:16:32	U	SMPL
CCV04 - 1	CCV04	27484	5.2476	-	7471B	PPB	13 Jan 2022 17:18:49	U	SMPL
CCB04 - 1	CCB04	-75	-0.0128	-	7471B	PPB	13 Jan 2022 17:21:06	U	SMPL