



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118219

Slope : 98.9

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/13/2022	16:16
2	CAL2	1	Water	NA	NA	20.6	7.00	01/13/2022	16:19
3	CAL3	1	Water	NA	NA	20.7	10.03	01/13/2022	16:20
4	ICV	1	Water	NA	NA	20.6	6.99	01/13/2022	16:22
5	CCV1	1	Water	NA	NA	20.4	2.02	01/13/2022	16:24
6	N1147-01	1	Water	NA	NA	20.4	6.89	01/13/2022	16:25
7	N1147-01DUP	1	Water	NA	NA	20.5	6.90	01/13/2022	16:26
8	N1147-03	1	Water	NA	NA	20.3	7.21	01/13/2022	16:30
9	CCV2	1	Water	NA	NA	20.6	12.03	01/13/2022	16:31

UP 118219

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHWN1147 WorkList ID : 156005 Department : Wet-Chemistry Date : 01-13-2022 16:00:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/13/2022	Water	N1147-01	pH	Cool 4 deg C	PSEG03	I11	DSN002	01/13/2022	9040C
01/13/2022	Water	N1147-03	pH	Cool 4 deg C	PSEG03	I11	RSW003	01/13/2022	9040C

Date/Time 01-13-22 16:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 01-13-22 17:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]