



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118224

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/13/2022	17:45
2	CAL2	1	Water	NA	NA	20.6	7.01	01/13/2022	17:46
3	CAL3	1	Water	NA	NA	20.6	10.03	01/13/2022	17:47
4	ICV	1	Water	NA	NA	20.5	6.98	01/13/2022	17:48
5	CCV1	1	Water	NA	NA	20.6	2.01	01/13/2022	17:49
6	N1112-02	1	Solid	20.03	20	20.1	8.42	01/13/2022	18:00
7	N1112-05	1	Solid	20.04	20	20.4	8.70	01/13/2022	18:01
8	N1121-03	1	Solid	20.02	20	20.2	7.74	01/13/2022	18:03
9	N1124-04	1	Solid	20.03	20	20.4	7.40	01/13/2022	18:05
10	N1130-02	1	Solid	20.02	20	20.2	8.33	01/13/2022	18:10
11	N1130-05	1	Solid	20.03	20	20.8	7.94	01/13/2022	18:13
12	N1136-02	1	Solid	20.02	20	20.5	8.70	01/13/2022	18:15
13	N1136-05	1	Solid	20.03	20	20.6	8.27	01/13/2022	18:19
14	N1136-05DUP	1	Solid	20.04	20	20.6	8.28	01/13/2022	18:20
15	CCV2	1	Water	NA	NA	20.7	12.03	01/13/2022	18:22

VB 11822-4

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1121			WorkList ID : 155987		Department : Wet-Chemistry		Date : 01-13-2022 10:02:47		
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2022	Solid	N1121-03	Corrosivity	Cool 4 deg C	ENVI27	H11	WASTE-CLASS-LAUNDR	01/12/2022	9045D
01/13/2022	Solid	N1124-04	Corrosivity	Cool 4 deg C	ENVI27	H11	WASTE-CLASS-BOILER-	01/13/2022	9045D
01/11/2022	Solid	N1112-02	Corrosivity	Cool 4 deg C	PSEG03	I11	TP-4S	01/11/2022	9045D
01/11/2022	Solid	N1112-05	Corrosivity	Cool 4 deg C	PSEG03	I11	TP-4D	01/11/2022	9045D
01/12/2022	Solid	N1130-02	Corrosivity	Cool 4 deg C	PSEG03	I11	TP-5S	01/12/2022	9045D
01/12/2022	Solid	N1130-05	Corrosivity	Cool 4 deg C	PSEG03	I11	TP-5D	01/12/2022	9045D
01/13/2022	Solid	N1136-02	Corrosivity	Cool 4 deg C	PSEG03	I21	TP-6S	01/13/2022	9045D
01/13/2022	Solid	N1136-05	Corrosivity	Cool 4 deg C	PSEG03	I21	TP-6D	01/13/2022	9045D

Date/Time 01-13-22 16:50
 Raw Sample Received by: JD (PWC)
 Raw Sample Relinquished by: [Signature]

Date/Time 01-13-22 18:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: JD (PWC)