

LB118226


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

39268



A= 0.0000e+000

B= 2.5572e-004

C= -4.2960e-002

Rho= 0.9997270

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.029	-0.029	53	0.000	53					
0.2	0.200	0.181	-0.019	874	0.0 %	874					-10
2.5	2.500	2.495	-0.005	9923	0.0 %	9923					0
5.0	5.000	5.172	0.172	20395	0.0 %	20395					3
7.5	7.500	7.383	-0.117	29041	0.0 %	29041					-2
10.0	10.000	9.999	-0.001	39268	0.0 %	39268					0

LB118226 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Jan 2022 10:55:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000	7470A	PPB	14 Jan 2022 11:00:13	S	Std
0.2 - 1	50.2	874	-	0.2000	7470A	PPB	14 Jan 2022 11:02:30	S	Std
2.5 - 1	52.5	9923	-	2.5000	7470A	PPB	14 Jan 2022 11:04:47	S	Std
5.0 - 1	55	20395	-	5.0000	7470A	PPB	14 Jan 2022 11:07:04	S	Std
7.5 - 1	57.5	29041	-	7.5000	7470A	PPB	14 Jan 2022 11:11:53	S	Std
10.0 - 1	516	39268	-	10.0000	7470A	PPB	14 Jan 2022 11:19:27	S	Std
ICV45 - 1	ICV45	15536	3.9299	-	7470A	PPB	14 Jan 2022 11:22:05	U	SMPL
ICB45 - 1	ICB45	-36	-0.0522	-	7470A	PPB	14 Jan 2022 11:24:20	U	SMPL
CCV05 - 1	CCV05	19768	5.0121	-	7470A	PPB	14 Jan 2022 11:26:35	U	SMPL
CCB05 - 1	CCB05	-39	-0.0529	-	7470A	PPB	14 Jan 2022 11:28:50	U	SMPL
CRA - 1	CRA	825	0.1680	-	7470A	PPB	14 Jan 2022 11:31:05	U	SMPL
HighStd - 1	HighStd	38947	9.9165	-	7470A	PPB	14 Jan 2022 11:33:20	U	SMPL
ChkStd - 1	ChkStd	26482	6.7290	-	7470A	PPB	14 Jan 2022 11:35:35	U	SMPL
PB142040BL - 1	PBW	-60	-0.0583	-	7470A	PPB	14 Jan 2022 11:37:52	U	SMPL
PB142040BS - 1	LCSW	13740	3.4706	-	7470A	PPB	14 Jan 2022 11:40:08	U	SMPL
N1098-03 - 1	14-G-WATER	44	-0.0317	-	7470A	PPB	14 Jan 2022 11:42:25	U	SMPL
N1098-03DUP - 1	14-G-WATERDUP	82	-0.0220	-	7470A	PPB	14 Jan 2022 11:44:41	U	SMPL
N1098-03MS - 1	14-G-WATERMS	14970	3.7851	-	7470A	PPB	14 Jan 2022 11:46:58	U	SMPL
N1098-03MSD - 1	14-G-WATERMSD	13571	3.4274	-	7470A	PPB	14 Jan 2022 11:49:15	U	SMPL
N1113-01 - 1	20220012-KEANSBURG-2-SITE-WATER	81	-0.0222	-	7470A	PPB	14 Jan 2022 11:51:30	U	SMPL
CCV06 - 1	CCV06	19280	4.8873	-	7470A	PPB	14 Jan 2022 11:53:45	U	SMPL
CCB06 - 1	CCB06	-49	-0.0555	-	7470A	PPB	14 Jan 2022 11:56:00	U	SMPL
N1113-02 - 1	20220013-KEANSBURG-3-SITE-WATER	182	0.0036	-	7470A	PPB	14 Jan 2022 11:58:15	U	SMPL
N1098-03LX5 - 1		9	-0.0407	-	7470A	PPB	14 Jan 2022 12:00:30	U	SMPL
N1098-03A - 1		14983	3.7885	-	7470A	PPB	14 Jan 2022 12:02:46	U	SMPL
CCV07 - 1	CCV07	19202	4.8673	-	7470A	PPB	14 Jan 2022 12:05:01	U	SMPL
CCB07 - 1	CCB07	-58	-0.0578	-	7470A	PPB	14 Jan 2022 12:07:19	U	SMPL