

LB118231


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

46526



A= 0.0000e+000

B= 2.1442e-004

C= -3.1425e-003

Rho= 0.9998433

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	6	0.000	6					
0.2	0.200	0.210	0.010	994	0.0 %	994					5
2.5	2.500	2.538	0.038	11853	0.0 %	11853					2
5.0	5.000	4.883	-0.117	22787	0.0 %	22787					-2
7.5	7.500	7.598	0.098	35449	0.0 %	35449					1
10.0	10.000	9.973	-0.027	46526	0.0 %	46526					0

LB118231 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jan 2022 12:27:11

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	6	-	0.00007471B	PPB	14 Jan 2022 12:28:50	S	Std
0.2 - 1	50.2	994	-	0.20007471B	PPB	14 Jan 2022 12:31:07	S	Std
2.5 - 1	52.5	11853	-	2.50007471B	PPB	14 Jan 2022 12:33:24	S	Std
5.0 - 1	55	22787	-	5.00007471B	PPB	14 Jan 2022 12:35:40	S	Std
7.5 - 1	57.5	35449	-	7.50007471B	PPB	14 Jan 2022 12:37:58	S	Std
10.0 - 1	50	46526	-	10.00007471B	PPB	14 Jan 2022 12:42:53	S	Std
ICV46 - 1	ICV46	19246	4.1236	-7471B	PPB	14 Jan 2022 12:45:38	U	SMPL
ICB46 - 1	ICB46	-39	-0.0115	-7471B	PPB	14 Jan 2022 12:47:53	U	SMPL
CCV08 - 1	CCV08	24980	5.3530	-7471B	PPB	14 Jan 2022 12:50:09	U	SMPL
CCB08 - 1	CCB08	-15	-0.0064	-7471B	PPB	14 Jan 2022 12:52:24	U	SMPL
CRA - 1	CRA	1023	0.2162	-7471B	PPB	14 Jan 2022 12:54:40	U	SMPL
HighStd - 1	HighStd	47876	10.2624	-7471B	PPB	14 Jan 2022 12:59:37	U	SMPL
ChkStd - 1	ChkStd	34084	7.3051	-7471B	PPB	14 Jan 2022 13:01:53	U	SMPL
PB142045BL - 1	PBS	-90	-0.0224	-7471B	PPB	14 Jan 2022 13:04:09	U	SMPL
PB142045BS - 1	LCSS	20425	4.3764	-7471B	PPB	14 Jan 2022 13:06:27	U	SMPL
N1124-03 - 1	WASTE-CLASS-BOILER-PIT	86745	18.5966	-7471B	PPB	14 Jan 2022 13:08:43	U	SMPL
N1130-01 - 1	TP-5S	6983	1.4941	-7471B	PPB	14 Jan 2022 13:11:00	U	SMPL
N1130-04 - 1	TP-5D	1116	0.2361	-7471B	PPB	14 Jan 2022 13:13:19	U	SMPL
N1136-01 - 1	TP-6S	14078	3.0154	-7471B	PPB	14 Jan 2022 13:15:37	U	SMPL
N1136-04 - 1	TP-6D	930	0.1963	-7471B	PPB	14 Jan 2022 13:17:52	U	SMPL
CCV09 - 1	CCV09	24076	5.1592	-7471B	PPB	14 Jan 2022 13:22:51	U	SMPL
CCB09 - 1	CCB09	-8	-0.0049	-7471B	PPB	14 Jan 2022 13:25:07	U	SMPL
N1136-04DUP - 1	TP-6DDUP	935	0.1973	-7471B	PPB	14 Jan 2022 13:27:22	U	SMPL
N1136-04MS - 1	TP-6DMS	20624	4.4190	-7471B	PPB	14 Jan 2022 13:32:07	U	SMPL
N1136-04MSD - 1	TP-6DMSD	18304	3.9216	-7471B	PPB	14 Jan 2022 13:34:22	U	SMPL
N1141-03 - 1	41862	372	0.0766	-7471B	PPB	14 Jan 2022 13:36:38	U	SMPL
N1141-04 - 1	41861	322	0.0659	-7471B	PPB	14 Jan 2022 13:38:54	U	SMPL
N1141-05 - 1	41172	450	0.0933	-7471B	PPB	14 Jan 2022 13:41:10	U	SMPL
N1141-06 - 1	101224	248	0.0500	-7471B	PPB	14 Jan 2022 13:43:27	U	SMPL
N1141-07 - 1	31272	304	0.0620	-7471B	PPB	14 Jan 2022 13:45:44	U	SMPL
N1143-01 - 1	WEST-BOILER-ROOM-SOUTH-SIDE-WALL	536	0.1118	-7471B	PPB	14 Jan 2022 13:48:01	U	SMPL
N1143-02 - 1	WEST-BOILER-ROOM-WEST-SIDE-WALL	1119	0.2368	-7471B	PPB	14 Jan 2022 13:50:17	U	SMPL
CCV10 - 1	CCV10	23077	4.9450	-7471B	PPB	14 Jan 2022 13:52:32	U	SMPL
CCB10 - 1	CCB10	-2	-0.0036	-7471B	PPB	14 Jan 2022 13:54:48	U	SMPL
N1143-03 - 1	WEST-BOILER-ROOM-BOTTOM	773	0.1626	-7471B	PPB	14 Jan 2022 13:57:04	U	SMPL
N1143-04 - 1	WEST-BOILER-ROOM-EAST-SIDE-WALL	143	0.0275	-7471B	PPB	14 Jan 2022 13:59:19	U	SMPL
N1143-05 - 1	WEST-BOILER-ROOM-NORTH-SIDE-WALL	746	0.1568	-7471B	PPB	14 Jan 2022 14:01:35	U	SMPL
N1144-01 - 1	RO-ETGI-352	21349	4.5745	-7471B	PPB	14 Jan 2022 14:03:50	U	SMPL
N1136-04LX5 - 1		169	0.0331	-7471B	PPB	14 Jan 2022 14:06:06	U	SMPL
N1136-04A - 1		19571	4.1932	-7471B	PPB	14 Jan 2022 14:08:23	U	SMPL
N1124-03DLX5 - 1		17029	3.6482	-7471B	PPB	14 Jan 2022 14:10:40	U	SMPL
CCV11 - 1	CCV11	23737	5.0865	-7471B	PPB	14 Jan 2022 14:12:56	U	SMPL
CCB11 - 1	CCB11	0	-0.0031	-7471B	PPB	14 Jan 2022 14:15:13	U	SMPL