

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

1/14/2022 14:39

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.791	0.0	0.072	
ICB1	0.226	0.0	0.001	
CCV1	244.418	0.0	0.184	
CCB1	0.203	0.0	0.001	
PB142036BL	0.230	0.0	0.001	
PB142036BS	96.069	0.0	0.073	
LOWPB142036	9.288	0.0	0.007	
HIGHPB142036	482.341	0.0	0.362	
N1113-01	0.774	0.0	0.001	
N1113-01DUP	0.790	0.0	0.001	
N1113-01MS	32.676	0.0	0.025	
N1113-01MSD	32.474	0.0	0.025	
N1113-02	0.231	0.0	0.001	
PB142035BL	0.201	0.0	0.001	
CCV2	246.021	0.0	0.185	
CCB2	0.234	0.0	0.001	
PB142035BS	95.182	0.0	0.072	
LOWPB142035	10.042	0.0	0.008	
HIGHPB142035	486.312	0.0	0.365	
N1112-04	1.830	0.0	0.002	
N1121-01	0.277	0.0	0.001	
N1124-03	2.225	0.0	0.002	
CCV3	247.598	0.0	0.186	
CCB3	0.150	0.0	0.001	
N1130-01	0.533	0.0	0.001	
N1130-04	0.201	0.0	0.001	
N1136-01	2.324	0.0	0.002	
N1136-04	0.267	0.0	0.001	
N1144-01	0.220	0.0	0.001	
N1112-01	0.300	0.0	0.001	
N1112-01DUP	0.135	0.0	0.001	
N1112-01MS	39.549	0.0	0.030	
N1112-01MSD	39.675	0.0	0.030	
CCV4	257.021	0.0	0.193	
CCB4	0.174	0.0	0.001	

N	35
Mean	69.285
SD	130.7789
CV%	188.75

Aquakem v. 7.2AQ1

Results from time period:

Fri Jan 14 13:29:32 2022

Fri Jan 14 14:36:49 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.5014	µg/l	1/14/2022 11:05:10	
5.OPPBCN	A	Total CN	P	5.1613	µg/l	1/14/2022 11:05:11	
10PPBCN	A	Total CN	P	9.7678	µg/l	1/14/2022 11:05:12	
50PPBCN	A	Total CN	P	48.3807	µg/l	1/14/2022 11:05:13	
100PPBCN	A	Total CN	P	99.1087	µg/l	1/14/2022 11:05:14	
250PPBCN	A	Total CN	P	253.4737	µg/l	1/14/2022 11:05:15	
500PPBCN	A	Total CN	P	498.6064	µg/l	1/14/2022 11:05:16	
ICV1	S	Total CN	P	94.7911	µg/l	1/14/2022 13:29:32	
ICB1	S	Total CN	P	0.226	µg/l	1/14/2022 13:29:34	
CCV1	S	Total CN	P	244.418	µg/l	1/14/2022 13:29:37	
CCB1	S	Total CN	P	0.2034	µg/l	1/14/2022 13:29:39	
PB142036BL	S	Total CN	P	0.2299	µg/l	1/14/2022 13:29:40	
PB142036BS	S	Total CN	P	96.0694	µg/l	1/14/2022 13:29:41	
LOWPB142036	S	Total CN	P	9.2876	µg/l	1/14/2022 13:29:42	
HIGHPB142036	S	Total CN	P	482.3407	µg/l	1/14/2022 13:37:07	
N1113-01	S	Total CN	P	0.7738	µg/l	1/14/2022 13:37:08	
N1113-01DUP	S	Total CN	P	0.7897	µg/l	1/14/2022 13:37:09	
N1113-01MS	S	Total CN	P	32.6756	µg/l	1/14/2022 13:37:10	
N1113-01MSD	S	Total CN	P	32.4742	µg/l	1/14/2022 13:37:11	
N1113-02	S	Total CN	P	0.2312	µg/l	1/14/2022 13:37:14	
PB142035BL	S	Total CN	P	0.2006	µg/l	1/14/2022 13:37:15	
CCV2	S	Total CN	P	246.0211	µg/l	1/14/2022 13:37:16	
CCB2	S	Total CN	P	0.2338	µg/l	1/14/2022 13:37:17	
PB142035BS	S	Total CN	P	95.182	µg/l	1/14/2022 13:44:40	
LOWPB142035	S	Total CN	P	10.0422	µg/l	1/14/2022 13:44:41	
HIGHPB142035	S	Total CN	P	486.3124	µg/l	1/14/2022 13:44:42	
N1112-04	S	Total CN	P	1.8297	µg/l	1/14/2022 13:44:49	
N1121-01	S	Total CN	P	0.2767	µg/l	1/14/2022 13:44:50	
N1124-03	S	Total CN	P	2.2251	µg/l	1/14/2022 13:51:47	
CCV3	S	Total CN	P	247.5984	µg/l	1/14/2022 13:51:48	
CCB3	S	Total CN	P	0.1497	µg/l	1/14/2022 13:51:49	
N1130-01	S	Total CN	P	0.5332	µg/l	1/14/2022 13:51:50	
N1130-04	S	Total CN	P	0.2006	µg/l	1/14/2022 13:51:51	
N1136-01	S	Total CN	P	2.3241	µg/l	1/14/2022 13:51:52	
N1136-04	S	Total CN	P	0.2665	µg/l	1/14/2022 13:51:53	
N1144-01	S	Total CN	P	0.2197	µg/l	1/14/2022 13:51:54	
N1112-01	S	Total CN	P	0.3	µg/l	1/14/2022 14:36:42	
N1112-01DUP	S	Total CN	P	0.1348	µg/l	1/14/2022 14:36:45	
N1112-01MS	S	Total CN	P	39.5488	µg/l	1/14/2022 14:36:46	
N1112-01MSD	S	Total CN	P	39.6752	µg/l	1/14/2022 14:36:47	
CCV4	S	Total CN	P	257.0208	µg/l	1/14/2022 14:36:48	
CCB4	S	Total CN	P	0.1745	µg/l	1/14/2022 14:36:49	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : Re

Instrument ID : Konelab

1/14/2022 11:06

Test Total CN


Accepted 1/14/2022 11:06

Factor 1333

Bias 0

Coeff. of det. 0.999914

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.5014	0.0000	-
2	5.0PPBCN	0.004	5.1613	5.0000	3.2
3	10PPBCN	0.008	9.7678	10.0000	-2.3
4	50PPBCN	0.037	48.3807	50.0000	-3.2
5	100PPBCN	0.075	99.1087	100.0000	-0.9
6	250PPBCN	0.191	253.4737	250.0000	1.4
7	500PPBCN	0.374	498.6064	500.0000	-0.3