


LB118238

7470A

INSTRUMENT ID: CV1

μ Abs.:

37608



Linear

A= 0.0000e+000

B= 2.6484e-004

C= -3.2073e-002

Rho= 0.9999145

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.020	-0.020	46	0.000	46					-
0.2	0.200	0.162	-0.038	734	0.0 %	734					-19
2.5	2.500	2.517	0.017	9624	0.0 %	9624					/
5.0	5.000	5.070	0.070	19264	0.0 %	19264					/
7.5	7.500	7.543	0.043	28601	0.0 %	28601					/
10.0	10.000	9.928	-0.072	37608	0.0 %	37608					-/

LB118238 INSTRUMENT ID :: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Jan 2022 14:50:08

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7470A	PPB	14 Jan 2022 14:51:04	S	Std
0.2 - 1	50.2	734	-	0.2000	7470A	PPB	14 Jan 2022 14:53:21	S	Std
2.5 - 1	52.5	9624	-	2.5000	7470A	PPB	14 Jan 2022 14:55:38	S	Std
5.0 - 1	55	19264	-	5.0000	7470A	PPB	14 Jan 2022 14:57:55	S	Std
7.5 - 1	57.5	28601	-	7.5000	7470A	PPB	14 Jan 2022 15:00:11	S	Std
10.0 - 1	514	37608	-	10.0000	7470A	PPB	14 Jan 2022 15:07:32	S	Std
ICV47 - 1	ICV47	14812	3.8908	-	7470A	PPB	14 Jan 2022 15:10:16	U	SMPL
ICB47 - 1	ICB47	-40	-0.0427	-	7470A	PPB	14 Jan 2022 15:12:31	U	SMPL
CCV12 - 1	CCV12	17970	4.7272	-	7470A	PPB	14 Jan 2022 15:14:46	U	SMPL
CCB12 - 1	CCB12	-32	-0.0405	-	7470A	PPB	14 Jan 2022 15:17:02	U	SMPL
CRA - 1	CRA	888	0.2031	-	7470A	PPB	14 Jan 2022 15:19:17	U	SMPL
HighStd - 1	HighStd	38890	10.2677	-	7470A	PPB	14 Jan 2022 15:21:32	U	SMPL
ChkStd - 1	ChkStd	25060	6.6049	-	7470A	PPB	14 Jan 2022 15:23:47	U	SMPL
PB142046BL - 1	PBW	-60	-0.0480	-	7470A	PPB	14 Jan 2022 15:26:03	U	SMPL
PB142046BS - 1	LCSW	15411	4.0494	-	7470A	PPB	14 Jan 2022 15:28:19	U	SMPL
N1112-02 - 1	TP-4S	-44	-0.0437	-	7470A	PPB	14 Jan 2022 15:30:36	U	SMPL
N1112-05 - 1	TP-4D	-4	-0.0331	-	7470A	PPB	14 Jan 2022 15:32:52	U	SMPL
N1121-03 - 1	WASTE-CLASS-LAUNDRY-PIT	-5	-0.0334	-	7470A	PPB	14 Jan 2022 15:35:09	U	SMPL
N1124-04 - 1	WASTE-CLASS-BOILER-PIT	20	-0.0268	-	7470A	PPB	14 Jan 2022 15:37:26	U	SMPL
N1130-02 - 1	TP-5S	26	-0.0252	-	7470A	PPB	14 Jan 2022 15:39:41	U	SMPL
CCV13 - 1	CCV13	18628	4.9014	-	7470A	PPB	14 Jan 2022 15:41:56	U	SMPL
CCB13 - 1	CCB13	-42	-0.0432	-	7470A	PPB	14 Jan 2022 15:44:12	U	SMPL
N1130-05 - 1	TP-5D	9	-0.0297	-	7470A	PPB	14 Jan 2022 15:46:27	U	SMPL
N1136-02 - 1	TP-6S	11	-0.0292	-	7470A	PPB	14 Jan 2022 15:48:43	U	SMPL
N1136-05 - 1	TP-6D	14	-0.0284	-	7470A	PPB	14 Jan 2022 15:50:58	U	SMPL
N1136-05DUP - 1	TP-6DDUP	-1	-0.0323	-	7470A	PPB	14 Jan 2022 15:53:14	U	SMPL
N1136-05MS - 1	TP-6DMS	16695	4.3895	-	7470A	PPB	14 Jan 2022 15:55:29	U	SMPL
N1136-05MSD - 1	TP-6DMSD	14695	3.8598	-	7470A	PPB	14 Jan 2022 15:57:46	U	SMPL
N1142-01 - 1	11221	424	0.0802	-	7470A	PPB	14 Jan 2022 16:00:02	U	SMPL
N1136-05LX5 - 1		-35	-0.0413	-	7470A	PPB	14 Jan 2022 16:02:19	U	SMPL
N1136-05A - 1		15551	4.0865	-	7470A	PPB	14 Jan 2022 16:04:36	U	SMPL
PB142009TB - 1	PB142009TB	-30	-0.0400	-	7470A	PPB	14 Jan 2022 16:06:51	U	SMPL
CCV14 - 1	CCV14	17820	4.6874	-	7470A	PPB	14 Jan 2022 16:09:07	U	SMPL
CCB14 - 1	CCB14	-43	-0.0435	-	7470A	PPB	14 Jan 2022 16:11:22	U	SMPL