



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118241

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/14/2022	16:22
2	CAL2	1	Water	NA	NA	20.3	7.00	01/14/2022	16:23
3	CAL3	1	Water	NA	NA	20.3	10.03	01/14/2022	16:25
4	ICV	1	Water	NA	NA	20.4	6.99	01/14/2022	16:30
5	CCV1	1	Water	NA	NA	20.4	2.01	01/14/2022	16:32
6	N1153-01	1	Solid	20.03	20	20.2	5.55	01/14/2022	16:35
7	N1153-01DUP	1	Solid	20.04	20	20.3	5.56	01/14/2022	16:36
8	CCV2	1	Water	NA	NA	20.4	12.03	01/14/2022	16:39

WB 11824)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : [H N1153 WorkList ID : 156044 Department : Wet-Chemistry Date : 01-14-2022 16:13:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/14/2022	Solid	N1153-01	pH	Cool 4 deg C	PSEG03	I11	31502	01/14/2022	9045D

Date/Time 01-14-22 16:20
 Raw Sample Received by: JD (WCC)
 Raw Sample Relinquished by: JT (SM)

Date/Time 01-14-22 14:35
 Raw Sample Received by: JT (SM)
 Raw Sample Relinquished by: JD (WCC)