



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/17/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 09:20
In Date: 01/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 15:25
Out Date: 01/17/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS- OVEN

QC:LB118247

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1169-01	BGLW7	1	1.15	8.70	9.85	9.15	92.0	
N1169-02	BGLW8	2	1.19	8.76	9.95	9.07	90.0	
N1169-03	BGLW9	3	1.19	8.72	9.91	9.06	90.3	
N1169-04	BGLX0	4	1.18	8.60	9.78	8.33	83.1	
N1169-05	BGLX1	5	1.13	8.45	9.58	8.15	83.1	
N1169-06	BGLX2	6	1.16	8.52	9.68	9.17	94.0	
N1169-07	BGLX3	7	1.19	8.55	9.74	9.03	91.7	
N1169-08	BGLX4	8	1.15	8.38	9.53	8.8	91.3	
N1169-09	BGLX5	9	1.15	8.82	9.97	8.94	88.3	
N1169-10	BGLX6	10	1.16	8.80	9.96	8.38	82.0	
N1169-11	BGLX8	11	1.19	8.56	9.75	8.72	88.0	
N1169-12	BGLX9	12	1.17	8.49	9.66	9.00	92.2	
N1169-13	BGLY0	13	1.18	8.49	9.67	8.55	86.8	
N1169-14	BGLY1	14	1.16	8.82	9.98	8.61	84.5	
N1169-15	BGLY2	15	1.19	8.69	9.88	8.67	86.1	
N1169-16	BGLY3	16	1.18	8.46	9.64	8.67	88.5	
N1169-17	BGLY4	17	1.17	8.56	9.73	8.62	87.0	
N1169-18	BGLY5	18	1.14	8.84	9.98	8.78	86.4	
N1169-19	BGLY6	19	1.12	8.67	9.79	8.25	82.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$