



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118249

Slope : 98.7

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/17/2022	11:42
2	CAL2	1	Water	NA	NA	20.6	7.00	01/17/2022	11:43
3	CAL3	1	Water	NA	NA	20.6	10.03	01/17/2022	11:45
4	ICV	1	Water	NA	NA	20.7	6.99	01/17/2022	11:47
5	CCV1	1	Water	NA	NA	20.6	2.01	01/17/2022	11:50
6	N1162-40	1	Water	NA	NA	20.1	6.91	01/17/2022	11:53
7	N1162-40DUP	1	Water	NA	NA	20.2	6.92	01/17/2022	11:54
8	CCV2	1	Water	NA	NA	20.6	12.02	01/17/2022	12:00

VB 118249

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1162 W WorkList ID : 156071 Department : Wet-Chemistry Date : 01-17-2022 11:35:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/14/2022	Water	N1162-40	pH	Cool 4 deg C	PSEG03	J11	1-5-TOTES	01/14/2022	9040C

Date/Time 01-17-22 11:40
Raw Sample Received by: JP WCC
Raw Sample Relinquished by: JP WCC

Date/Time 01-17-22 12:30
Raw Sample Received by: JP WCC
Raw Sample Relinquished by: JP WCC