


LB118264

7470A

INSTRUMENT ID: CV1

 μ Abs.:

41025



Linear

A= 0.0000e+000

B= 2.4607e-004

C= 3.4230e-002

Rho= 0.9996735

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	-85	0.000	-85					-
0.2	0.200	0.224	0.024	771	0.0 %	771					12
2.5	2.500	2.519	0.019	10096	0.0 %	10096					1
5.0	5.000	5.004	0.004	20195	0.0 %	20195					0
7.5	7.500	7.311	-0.189	29574	0.0 %	29574					-3
10.0	10.000	10.129	0.129	41025	0.0 %	41025					1

LB118264 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jan 2022 10:02:27

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-85	-	0.00007470A	PPB	18 Jan 2022 10:25:07	S	Std
0.2 - 1	52	771	-	0.20007470A	PPB	18 Jan 2022 10:27:24	S	Std
2.5 - 1	52.5	10096	-	2.50007470A	PPB	18 Jan 2022 10:29:41	S	Std
5.0 - 1	53	20195	-	5.00007470A	PPB	18 Jan 2022 10:31:58	S	Std
7.5 - 1	53.5	29574	-	7.50007470A	PPB	18 Jan 2022 10:34:15	S	Std
10.0 - 1	54	41025	-	10.00007470A	PPB	18 Jan 2022 10:39:07	S	Std
ICV49 - 1	ICV49	17275	4.2851	-7470A	PPB	18 Jan 2022 10:44:19	U	SMPL
ICB49 - 1	ICB49	-59	0.0197	-7470A	PPB	18 Jan 2022 10:46:34	U	SMPL
CCV18 - 1	CCV18	19023	4.7152	-7470A	PPB	18 Jan 2022 10:48:49	U	SMPL
CCB18 - 1	CCB18	-64	0.0185	-7470A	PPB	18 Jan 2022 10:51:04	U	SMPL
CRA - 1	CRA	776	0.2252	-7470A	PPB	18 Jan 2022 10:53:19	U	SMPL
HighStd - 1	HighStd	40754	10.0625	-7470A	PPB	18 Jan 2022 10:55:34	U	SMPL
ChkStd - 1	ChkStd	28608	7.0737	-7470A	PPB	18 Jan 2022 10:57:50	U	SMPL
PB142081BL - 1	PBW	-82	0.0141	-7470A	PPB	18 Jan 2022 11:00:06	U	SMPL
PB142081BS - 1	LCSW	15239	3.7841	-7470A	PPB	18 Jan 2022 11:02:23	U	SMPL
N1153-01 - 1	31502	-39	0.0246	-7470A	PPB	18 Jan 2022 11:04:39	U	SMPL
N1153-01DUP - 1	31502DUP	13	0.0374	-7470A	PPB	18 Jan 2022 11:06:55	U	SMPL
N1153-01MS - 1	31502MS	14720	3.6564	-7470A	PPB	18 Jan 2022 11:09:12	U	SMPL
N1153-01MSD - 1	31502MSD	13748	3.4172	-7470A	PPB	18 Jan 2022 11:11:29	U	SMPL
PB142082BL - 1	PBW	-42	0.0239	-7470A	PPB	18 Jan 2022 11:13:44	U	SMPL
CCV19 - 1	CCV19	18235	4.5213	-7470A	PPB	18 Jan 2022 11:15:59	U	SMPL
CCB19 - 1	CCB19	-62	0.0190	-7470A	PPB	18 Jan 2022 11:18:14	U	SMPL
PB142082BS - 1	LCSW	13919	3.4593	-7470A	PPB	18 Jan 2022 11:20:30	U	SMPL
N1154-03 - 1	20220011-KEANSBURG-1-SITE-WATER	5	0.0355	-7470A	PPB	18 Jan 2022 11:22:46	U	SMPL
N1154-04 - 1	20220010-FIELD-BLANK	0	0.0342	-7470A	PPB	18 Jan 2022 11:25:01	U	SMPL
N1154-05 - 1	20220031-KEANSBURG-2-MODIFIED-ELUTRIATE	80	0.0539	-7470A	PPB	18 Jan 2022 11:27:17	U	SMPL
N1154-07 - 1	20220032-KEANSBURG-3-MODIFIED-ELUTRIATE	40	0.0441	-7470A	PPB	18 Jan 2022 11:29:33	U	SMPL
N1162-40 - 1	1-5-TOTES	189	0.0807	-7470A	PPB	18 Jan 2022 11:31:49	U	SMPL
N1162-40DUP - 1	1-5-TOTESDUP	216	0.0874	-7470A	PPB	18 Jan 2022 11:34:05	U	SMPL
N1162-40MS - 1	1-5-TOTESMS	14038	3.4885	-7470A	PPB	18 Jan 2022 11:36:23	U	SMPL
N1162-40MSD - 1	1-5-TOTESMSD	15152	3.7627	-7470A	PPB	18 Jan 2022 11:41:12	U	SMPL
PB142083BL - 1	PBW	-59	0.0197	-7470A	PPB	18 Jan 2022 11:43:27	U	SMPL
CCV20 - 1	CCV20	20043	4.9662	-7470A	PPB	18 Jan 2022 11:45:43	U	SMPL
CCB20 - 1	CCB20	-58	0.0200	-7470A	PPB	18 Jan 2022 11:47:58	U	SMPL
PB142083BS - 1	LCSW	14396	3.5766	-7470A	PPB	18 Jan 2022 11:50:14	U	SMPL
N1152-02 - 1	VACLOT-TOPSOIL-01	30	0.0416	-7470A	PPB	18 Jan 2022 11:52:30	U	SMPL
N1154-09 - 1	20220028-KEANSBURG-2-COMP-SPLP-LEACHATE	134	0.0672	-7470A	PPB	18 Jan 2022 11:54:45	U	SMPL
N1154-09DUP - 1	20220028-KEANSBURG-2-COMP-SPLP-LEACHATEDUP	42	0.0446	-7470A	PPB	18 Jan 2022 11:57:01	U	SMPL
N1154-09MS - 1	20220028-KEANSBURG-2-COMP-SPLP-LEACHATEMS	13813	3.4332	-7470A	PPB	18 Jan 2022 12:01:34	U	SMPL
N1154-09MSD - 1	20220028-KEANSBURG-2-COMP-SPLP-LEACHATEMSD	13571	3.3736	-7470A	PPB	18 Jan 2022 12:03:51	U	SMPL
N1154-10 - 1	20220029-KEANSBURG-3-COMP-SPLP-LEACHATE	-11	0.0315	-7470A	PPB	18 Jan 2022 12:06:07	U	SMPL
PB142055TB - 1	PB142055TB	-2	0.0337	-7470A	PPB	18 Jan 2022 12:08:25	U	SMPL
PB142020TB - 1	PB142020TB	11	0.0369	-7470A	PPB	18 Jan 2022 12:10:40	U	SMPL
CCV21 - 1	CCV21	20106	4.9817	-7470A	PPB	18 Jan 2022 12:12:56	U	SMPL
CCB21 - 1	CCB21	-35	0.0256	-7470A	PPB	18 Jan 2022 12:15:11	U	SMPL
N1153-01LX5 - 1		-9	0.0320	-7470A	PPB	18 Jan 2022 12:17:27	U	SMPL
N1153-01A - 1		15895	3.9455	-7470A	PPB	18 Jan 2022 12:19:42	U	SMPL
N1162-40LX5 - 1		-4	0.0332	-7470A	PPB	18 Jan 2022 12:21:58	U	SMPL
N1162-40A - 1		17539	4.3500	-7470A	PPB	18 Jan 2022 12:24:13	U	SMPL
N1154-09LX5 - 1		-34	0.0259	-7470A	PPB	18 Jan 2022 12:26:29	U	SMPL
N1154-09A - 1		17023	4.2230	-7470A	PPB	18 Jan 2022 12:28:46	U	SMPL
CCV22 - 1	CCV22	19630	4.8645	-7470A	PPB	18 Jan 2022 12:31:02	U	SMPL
CCB22 - 1	CCB22	-74	0.0160	-7470A	PPB			