


LB118272

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.8524e-004

C= 3.2352e-002

Rho= 0.9998558

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10/11
0.0	0.000	0.038	0.038	29	0.000	29					19
0.2	0.200	0.238	0.038	1109	0.0 %	1109					-5
2.5	2.500	2.376	-0.124	12650	0.0 %	12650					1
5.0	5.000	5.059	0.059	27135	0.0 %	27135					0
7.5	7.500	7.466	-0.034	40127	0.0 %	40127					0
10.0	10.000	10.024	0.024	53940	0.0 %	53940					

LB118272 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Jan 2022 14:10:35

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.00007471B	PPB		18 Jan 2022 14:13:30	S	Std
0.2 - 1	50.2	1109	-	0.20007471B	PPB		18 Jan 2022 14:15:47	S	Std
2.5 - 1	52.5	12650	-	2.50007471B	PPB		18 Jan 2022 14:18:05	S	Std
5.0 - 1	35	27135	-	5.00007471B	PPB		18 Jan 2022 14:20:22	S	Std
7.5 - 1	37.5	40127	-	7.50007471B	PPB		18 Jan 2022 14:22:38	S	Std
10.0 - 1	38	53940	-	10.00007471B	PPB		18 Jan 2022 14:24:54	S	Std
ICV50 - 1	ICV50	20326	3.7976	-7471B	PPB		18 Jan 2022 14:27:47	U	SMPL
ICB50 - 1	ICB50	-42	0.0246	-7471B	PPB		18 Jan 2022 14:30:02	U	SMPL
CCV23 - 1	CCV23	26383	4.9196	-7471B	PPB		18 Jan 2022 14:32:17	U	SMPL
CCB23 - 1	CCB23	-56	0.0220	-7471B	PPB		18 Jan 2022 14:34:32	U	SMPL
CRA - 1	CRA	1037	0.2244	-7471B	PPB		18 Jan 2022 14:36:48	U	SMPL
HighStd - 1	HighStd	52341	9.7281	-7471B	PPB		18 Jan 2022 14:39:03	U	SMPL
ChkStd - 1	ChkStd	36411	6.7772	-7471B	PPB		18 Jan 2022 14:41:18	U	SMPL
PB142103BL - 1	PBS	-72	0.0190	-7471B	PPB		18 Jan 2022 14:43:36	U	SMPL
PB142103BS - 1	LCSS	18867	3.5273	-7471B	PPB		18 Jan 2022 14:45:52	U	SMPL
N1154-01 - 1	20220028-KEANSBURG-2-COMP	20007	3.7385	-7471B	PPB		18 Jan 2022 14:48:08	U	SMPL
N1154-02 - 1	20220029-KEANSBURG-3-COMP	12003	2.2558	-7471B	PPB		18 Jan 2022 14:50:25	U	SMPL
N1172-01 - 1	EME-TS-3	2435	0.4834	-7471B	PPB		18 Jan 2022 14:52:42	U	SMPL
N1172-02 - 1	EME-TS-4	2152	0.4310	-7471B	PPB		18 Jan 2022 14:54:59	U	SMPL
N1172-02DUP - 1	EME-TS-4DUP	2360	0.4695	-7471B	PPB		18 Jan 2022 14:57:14	U	SMPL
CCV24 - 1	CCV24	26163	4.8789	-7471B	PPB		18 Jan 2022 14:59:30	U	SMPL
CCB24 - 1	CCB24	-42	0.0246	-7471B	PPB		18 Jan 2022 15:01:45	U	SMPL
N1172-02MS - 1	EME-TS-4MS	23346	4.3570	-7471B	PPB		18 Jan 2022 15:04:01	U	SMPL
N1172-02MSD - 1	EME-TS-4MSD	22946	4.2829	-7471B	PPB		18 Jan 2022 15:06:17	U	SMPL
N1172-02LX5 - 1		355	0.0981	-7471B	PPB		18 Jan 2022 15:08:32	U	SMPL
N1172-02A - 1		24578	4.5852	-7471B	PPB		18 Jan 2022 15:10:48	U	SMPL
CCV25 - 1	CCV25	26275	4.8996	-7471B	PPB		18 Jan 2022 15:13:03	U	SMPL
CCB25 - 1	CCB25	-54	0.0223	-7471B	PPB		18 Jan 2022 15:15:20	U	SMPL