



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118278

Slope : 98.7

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/18/2022	16:45
2	CAL2	1	Water	NA	NA	20.6	7.00	01/18/2022	16:46
3	CAL3	1	Water	NA	NA	20.6	10.03	01/18/2022	16:47
4	ICV	1	Water	NA	NA	20.7	6.99	01/18/2022	16:49
5	CCV1	1	Water	NA	NA	20.6	2.01	01/18/2022	16:50
6	N1178-03	1	Water	NA	NA	20.6	7.43	01/19/2022	16:52
7	N1178-03DUP	1	Water	NA	NA	20.7	7.44	01/18/2022	16:53
8	N1178-05	1	Water	NA	NA	20.6	7.53	01/18/2022	16:55
9	N1180-01	1	Water	NA	NA	20.5	7.31	01/18/2022	16:56
10	N1187-01	1	Water	NA	NA	20.1	3.82	01/18/2022	16:59
11	N1187-02	1	Water	NA	NA	20.2	5.88	01/18/2022	17:00
12	N1187-03	1	Water	NA	NA	20.3	5.65	01/18/2022	17:02
13	N1187-04	1	Water	NA	NA	20.1	6.78	01/19/2022	17:03
14	N1187-06	1	Water	NA	NA	20.2	5.41	01/18/2022	17:05
15	N1187-07	1	Water	NA	NA	20.4	6.31	01/18/2022	17:06
16	CCV2	1	Water	NA	NA	20.6	12.03	01/18/2022	17:10
17	N1187-08	1	Water	NA	NA	20.4	7.34	01/18/2022	17:13
18	N1187-09	1	Water	NA	NA	20.3	7.52	01/18/2022	17:15
19	CCV3	1	Water	NA	NA	20.6	2.01	01/18/2022	17:19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1181 S

WorkList ID : 156111

Department : Wet-Chemistry

Date : 01-18-2022 16:40:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/18/2022	Water	N1178-03	✓ pH	Cool 4 deg C	PSEG03	J11	AUD-1244	01/18/2022	9040C
01/18/2022	Water	N1178-05	✓ pH	Cool 4 deg C	PSEG03	J11	AUD-1113	01/18/2022	9040C
01/18/2022	Water	N1180-01	✓ pH	Cool 4 deg C	PSEG03	J11	31522	01/18/2022	9040C
01/18/2022	Water	N1187-01	✓ pH	Cool 4 deg C	PSEG03	J11	20904	01/18/2022	9040C
01/18/2022	Water	N1187-02	✓ pH	Cool 4 deg C	PSEG03	J11	20963	01/18/2022	9040C
01/18/2022	Water	N1187-03	✓ pH	Cool 4 deg C	PSEG03	J11	20988	01/18/2022	9040C
01/18/2022	Water	N1187-04	✓ pH	Cool 4 deg C	PSEG03	J11	31298	01/18/2022	9040C
01/18/2022	Water	N1187-06	✓ pH	Cool 4 deg C	PSEG03	J11	31282	01/18/2022	9040C
01/18/2022	Water	N1187-07	✓ pH	Cool 4 deg C	PSEG03	J11	20970	01/18/2022	9040C
01/18/2022	Water	N1187-08	✓ pH	Cool 4 deg C	PSEG03	J11	31338	01/18/2022	9040C
01/18/2022	Water	N1187-09	✓ pH	Cool 4 deg C	PSEG03	J11	600492261	01/18/2022	9040C

Date/Time

01-18-22

16:42

Raw Sample Received by:

JA (WC)

Raw Sample Relinquished by:

JA (WC)

Date/Time

01-18-22

18:30

Raw Sample Received by:

JA (WC)

Raw Sample Relinquished by:

JA (WC)

VP 118278