



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118282

Slope : 98.7

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/18/2022	16:45
2	CAL2	1	Water	NA	NA	20.6	7.00	01/18/2022	16:46
3	CAL3	1	Water	NA	NA	20.6	10.03	01/18/2022	16:47
4	ICV	1	Water	NA	NA	20.7	6.99	01/18/2022	16:49
5	CCV1	1	Water	NA	NA	20.4	2.01	01/18/2022	18:10
6	N1186-01	1	Water	NA	NA	20.1	5.03	01/19/2022	18:11
7	N1186-01DUP	1	Water	NA	NA	20.1	5.04	01/18/2022	18:12
8	CCV2	1	Water	NA	NA	20.6	12.03	01/18/2022	18:15

VP 118282

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1186 WorkList ID : 156112 Department : Wet-Chemistry Date : 01-18-2022 16:41:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/18/2022	Water	N1186-01	Corrosivity	Cool 4 deg C	PSEG03	J11	17103	01/18/2022	9040C

Date/Time 01.18.22 17:15
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: J.C (SW)

Date/Time 01.18.22 18:30
Raw Sample Received by: J.C (SW)
Raw Sample Relinquished by: JP (WC)