



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118283

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/19/2022	09:15
2	N1141-01		1	8	45.00	113.0	113.0	01/19/2022	09:40
3	N1141-01DUP		1	8	48.00	118.4	118.4	01/19/2022	10:05
4	N1142-01		1	9	52.00	125.6	125.6	01/19/2022	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118283

WorkList Name : fp11422 WorkList ID : 156024 Department : Wet-Chemistry Date : 01-14-2022 09:19:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/20/2022	Solid	N1141-01	✓ Flash Point	Cool 4 deg C	PSEG03	I11	30650	01/13/2022	1010A
01/20/2022	Solid	N1142-01	✗ Flash Point	Cool 4 deg C	PSEG03	I11	11221	01/13/2022	1010A

Date/Time 01/19/22 08:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CETA (ab)

Date/Time 01/19/22 10:40
 Raw Sample Received by: RJ CETA (ab)
 Raw Sample Relinquished by: KS (WC)