



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118286

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/19/2022	09:15
2	N1186-08		1	9	45.00	113.0	113.0	01/19/2022	10:55
3	N1186-08DUP		1	9	47.00	116.6	116.6	01/19/2022	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP11922 WorkList ID : 156132 Department : Wet-Chemistry Date : 01-19-2022 10:48:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/21/2022	Solid	N1186-08	Flash Point	Cool 4 deg C	PSEG03	J11	17103	01/18/2022	1010B

Date/Time 01/19/22 10:55
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: RJ (EXT-126)

Date/Time 01/19/22 11:20
 Raw Sample Received by: RJ (EXT-126)
 Raw Sample Relinquished by: KS(WC)