



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118287

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/18/2022	10:05
2	N1134-01		1	10	100.00	>212.0	>212.0	01/18/2022	10:35
3	N1134-01DUP		1	10	100.00	>212.0	>212.0	01/18/2022	10:58
4	N1162-40		1	12	100.00	>212.0	>212.0	01/18/2022	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118287

WorkList Name : fp11422-1 WorkList ID : 156025 Department : Wet-Chemistry Date : 01-14-2022 09:20:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/26/2022	Water	N1134-01	Flash Point	Cool 4 deg C	TETR16	I11	MC0J40	01/12/2022	1010B
01/19/2022	Water	N1162-40	Flash Point	Cool 4 deg C	PSEG03	J11	1-5-TOTES	01/14/2022	1010B

Date/Time 01/18/22 09:30
Raw Sample Received by: KS (wc)
Raw Sample Relinquished by: RS c EAT-1066

Date/Time 01/18/22 11:25
Raw Sample Received by: RS c EAT-1066
Raw Sample Relinquished by: KS (wc)