



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118288

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	01/19/2022	12:45
2	N1178-03		1	10	100.00	>212.0	>212.0	01/19/2022	13:10
3	N1178-03DUP		1	10	100.00	>212.0	>212.0	01/19/2022	13:35
4	N1178-05		1	15	100.00	>212.0	>212.0	01/19/2022	13:58
5	N1180-01		1	12	100.00	>212.0	>212.0	01/19/2022	14:20
6	N1187-05		1	13	100.00	>212.0	>212.0	01/19/2022	14:45
7	N1187-06		1	14	100.00	>212.0	>212.0	01/19/2022	15:05
8	N1187-07		1	15	100.00	>212.0	>212.0	01/19/2022	15:35
9	N1187-08		1	14	100.00	>212.0	>212.0	01/19/2022	15:55
10	N1187-09		1	15	100.00	>212.0	>212.0	01/19/2022	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118288

WorkList Name : FP11922-1

WorkList ID : 156133

Department : Wet-Chemistry

Date : 01-19-2022 12:19:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/31/2022	Water	N1178-03	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1244	01/18/2022	1010B
01/31/2022	Water	N1178-05	Flash Point	Cool 4 deg C	PSEG03	J11	AUD-1113	01/18/2022	1010B
01/21/2022	Water	N1180-01	Flash Point	Cool 4 deg C	PSEG03	J11	31522	01/18/2022	1010B
01/25/2022	Water	N1187-05	Flash Point	Cool 4 deg C	PSEG03	J11	COMP-20904-20963-209	01/18/2022	1010B
01/25/2022	Water	N1187-06	Flash Point	Cool 4 deg C	PSEG03	J11	31282	01/18/2022	1010B
01/25/2022	Water	N1187-07	Flash Point	Cool 4 deg C	PSEG03	J11	20970	01/18/2022	1010B
01/25/2022	Water	N1187-08	Flash Point	Cool 4 deg C	PSEG03	J11	31338	01/18/2022	1010B
01/25/2022	Water	N1187-09	Flash Point	Cool 4 deg C	PSEG03	J11	600492261	01/18/2022	1010B

Date/Time 01/19/22 12:20
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CEX-106

Date/Time 01/19/22 16:25
 Raw Sample Received by: RJ CEX-106
 Raw Sample Relinquished by: KS (WC)