

LB118291


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44218



A= 0.0000e+000

B= 2.2773e-004 *slope*C= 2.8809e-003 *= intercept*

Rho= 0.9998510

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>9/21/11</i>
0.00	0.000	-0.002	-0.002	-23	0.000	-23					
0.05	0.050					205	0				
0.20	0.200	0.185	-0.015	800	0.0 %	800					
2.50	2.500	2.573	0.073	11286	0.0 %	11286					
5.00	5.000	4.987	-0.013	21886	0.0 %	21886					
7.50	7.500	7.385	-0.115	32414	0.0 %	32414					
10.0	10.000	10.073	0.073	44218	0.0 %	44218					

103021

LB118291

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	-23	-		0 SFAM01.1	PPB	1/19/2022 15:05	Std
	0.2 S01	800	-		0.2 SFAM01.1	PPB	1/19/2022 15:07	Std
	2.5 S02	11286	-		2.5 SFAM01.1	PPB	1/19/2022 15:10	Std
	5 S03	21886	-		5 SFAM01.1	PPB	1/19/2022 15:12	Std
	7.5 S04	32414	-		7.5 SFAM01.1	PPB	1/19/2022 15:14	Std
	10 S05	44218	-		10 SFAM01.1	PPB	1/19/2022 15:17	Std
ICV010	ICV010	17869	4.0722 -		SFAM01.1	PPB	1/19/2022 15:19	SMPL
ICB010	ICB010	-7	0.0013 -		SFAM01.1	PPB	1/19/2022 15:22	SMPL
CCV028	CCV028	22200	5.0585 -		SFAM01.1	PPB	1/19/2022 15:24	SMPL
CCB028	CCB028	-5	0.0017 -		SFAM01.1	PPB	1/19/2022 15:26	SMPL
PB142131BL	PBS131	17	0.0068 -		SFAM01.1	PPB	1/19/2022 15:28	SMPL
N1135-19	X3883	18076	4.1193 -		SFAM01.1	PPB	1/19/2022 15:33	SMPL
N1135-20	X3883D	18415	4.1965 -		SFAM01.1	PPB	1/19/2022 15:35	SMPL
N1135-21	X3883S	28747	6.5495 -		SFAM01.1	PPB	1/19/2022 15:42	SMPL
CCV029	CCV029	22369	5.097 -		SFAM01.1	PPB	1/19/2022 15:44	SMPL
CCB029	CCB029	-22	-0.0021 -		SFAM01.1	PPB	1/19/2022 15:46	SMPL