



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118293

Slope : 99.3

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/19/2022	16:05
2	CAL2	1	Water	NA	NA	20.3	7.00	01/19/2022	16:06
3	CAL3	1	Water	NA	NA	20.3	10.03	01/19/2022	16:07
4	ICV	1	Water	NA	NA	20.4	6.99	01/19/2022	16:10
5	CCV1	1	Water	NA	NA	20.4	2.01	01/19/2022	16:11
6	N1186-08	1	Solid	20.03	20	20.3	5.03	01/19/2022	16:12
7	N1188-02	1	Solid	20.04	20	20.5	6.72	01/19/2022	16:15
8	N1188-02DUP	1	Solid	20.05	20	20.6	6.73	01/19/2022	16:16
9	CCV2	1	Water	NA	NA	20.4	12.03	01/19/2022	16:19

WB 118293

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N5188 WorkList ID : 156141 Department : Wet-Chemistry Date : 01-19-2022 13:50:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/18/2022	Solid	N1186-08	Corrosivity	Cool 4 deg C	PSEG03	17103	17103	01/18/2022	9045D
01/19/2022	Solid	N1188-02	Corrosivity	Cool 4 deg C	PSEG03	J11	72-11940	01/19/2022	9045D

Date/Time 01.19.22 16:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 01.19.22 17:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]