

LB118297


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39827



A= 0.0000e+000

B= 2.5380e-004

C= -2.7170e-002

Rho= 0.9996618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/dl
0.0	0.000	-0.021	-0.021	26	0.000	26					1
0.2	0.200	0.179	-0.021	812	0.0 %	812					11
2.5	2.500	2.574	0.074	10251	0.0 %	10251					3
5.0	5.000	5.077	0.077	20112	0.0 %	20112					2
7.5	7.500	7.309	-0.191	28907	0.0 %	28907					3
10.0	10.000	10.081	0.081	39827	0.0 %	39827					1

LB118297 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jan 2022 09:28:17

Sample ID	Extended ID	μ Abs	Conc	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.00007470A	PPB		20 Jan 2022 09:32:07	S	Std
0.2 - 1	50.2	812	-	0.20007470A	PPB		20 Jan 2022 09:34:24	S	Std
2.5 - 1	52.7	10251	-	2.50007470A	PPB		20 Jan 2022 09:39:13	S	Std
5.0 - 1	55	20112	-	5.00007470A	PPB		20 Jan 2022 09:41:30	S	Std
7.5 - 1	59.5	28907	-	7.50007470A	PPB		20 Jan 2022 09:43:46	S	Std
10.0 - 1	511	39827	-	10.00007470A	PPB		20 Jan 2022 09:46:02	S	Std
ICV51 - 1	ICV51	15773	3.9759	-7470A	PPB		20 Jan 2022 09:49:06	U	SMPL
ICB51 - 1	ICB51	-28	-0.0343	-7470A	PPB		20 Jan 2022 09:51:21	U	SMPL
CCV30 - 1	CCV30	19983	5.0444	-7470A	PPB		20 Jan 2022 09:53:35	U	SMPL
CCB30 - 1	CCB30	-46	-0.0388	-7470A	PPB		20 Jan 2022 09:55:51	U	SMPL
CRA - 1	CRA	745	0.1619	-7470A	PPB		20 Jan 2022 09:58:05	U	SMPL
HighStd - 1	HighStd	39412	9.9754	-7470A	PPB		20 Jan 2022 10:00:21	U	SMPL
ChkStd - 1	ChkStd	26506	6.6999	-7470A	PPB		20 Jan 2022 10:02:36	U	SMPL
PB142127BL - 1	PBW	3	-0.0264	-7470A	PPB		20 Jan 2022 10:07:24	U	SMPL
PB142127BS - 1	LCSW	16516	4.1645	-7470A	PPB		20 Jan 2022 10:09:40	U	SMPL
N1176-02 - 1	SB-9	-21	-0.0325	-7470A	PPB		20 Jan 2022 10:11:56	U	SMPL
N1176-02DUP - 1	SB-9DUP	42	-0.0165	-7470A	PPB		20 Jan 2022 10:14:13	U	SMPL
N1176-02MS - 1	SB-9MS	15313	3.8592	-7470A	PPB		20 Jan 2022 10:16:29	U	SMPL
N1176-02MSD - 1	SB-9MSD	15973	4.0267	-7470A	PPB		20 Jan 2022 10:18:46	U	SMPL
N1178-01 - 1	AUD-1128	-33	-0.0355	-7470A	PPB		20 Jan 2022 10:21:01	U	SMPL
CCV31 - 1	CCV31	20780	5.2467	-7470A	PPB		20 Jan 2022 10:23:17	U	SMPL
CCB31 - 1	CCB31	-39	-0.0371	-7470A	PPB		20 Jan 2022 10:25:32	U	SMPL
N1186-02 - 1	COMP-1	13	-0.0239	-7470A	PPB		20 Jan 2022 10:27:48	U	SMPL
N1186-03 - 1	17109	33	-0.0188	-7470A	PPB		20 Jan 2022 10:30:03	U	SMPL
N1186-04 - 1	17104	36	-0.0180	-7470A	PPB		20 Jan 2022 10:32:18	U	SMPL
N1186-05 - 1	17101	28	-0.0201	-7470A	PPB		20 Jan 2022 10:34:34	U	SMPL
PB142128BL - 1	PBW	26	-0.0206	-7470A	PPB		20 Jan 2022 10:36:49	U	SMPL
PB142128BS - 1	LCSW	14763	3.7196	-7470A	PPB		20 Jan 2022 10:39:05	U	SMPL
N1178-03 - 1	AUD-1244	-2	-0.0277	-7470A	PPB		20 Jan 2022 10:41:22	U	SMPL
N1178-03DUP - 1	AUD-1244DUP	39	-0.0173	-7470A	PPB		20 Jan 2022 10:43:39	U	SMPL
N1178-03MS - 1	AUD-1244MS	16395	4.1338	-7470A	PPB		20 Jan 2022 10:45:56	U	SMPL
N1178-03MSD - 1	AUD-1244MSD	17487	4.4109	-7470A	PPB		20 Jan 2022 10:48:12	U	SMPL
CCV32 - 1	CCV32	20590	5.1985	-7470A	PPB		20 Jan 2022 10:50:27	U	SMPL
CCB32 - 1	CCB32	-15	-0.0310	-7470A	PPB		20 Jan 2022 10:52:43	U	SMPL
N1178-05 - 1	AUD-1113	56	-0.0130	-7470A	PPB		20 Jan 2022 10:54:58	U	SMPL
N1186-01 - 1	17103	19	-0.0223	-7470A	PPB		20 Jan 2022 10:57:14	U	SMPL
PB142129BL - 1	PBW	32	-0.0190	-7470A	PPB		20 Jan 2022 10:59:29	U	SMPL
PB142129BS - 1	LCSW	17968	4.5330	-7470A	PPB		20 Jan 2022 11:01:45	U	SMPL
N1154-06 - 1	20220031-KEANSBURG-2-MODIFIED-ELUTRIATE	30	-0.0196	-7470A	PPB		20 Jan 2022 11:04:00	U	SMPL
N1154-08 - 1	20220032-KEANSBURG-3-MODIFIED-ELUTRIATE	59	-0.0122	-7470A	PPB		20 Jan 2022 11:06:17	U	SMPL
N1187-08 - 1	31338	103684	26.2873	-7470A	PPB		20 Jan 2022 11:08:33	U	SMPL
N1187-09 - 1	600492261	53	-0.0137	-7470A	PPB		20 Jan 2022 11:13:30	U	SMPL
N1187-09DUP - 1	600492261DUP	58	-0.0125	-7470A	PPB		20 Jan 2022 11:15:47	U	SMPL
N1187-09MS - 1	600492261MS	13536	3.4082	-7470A	PPB		20 Jan 2022 11:18:03	U	SMPL
CCV33 - 1	CCV33	19756	4.9868	-7470A	PPB		20 Jan 2022 11:20:18	U	SMPL
CCB33 - 1	CCB33	-18	-0.0317	-7470A	PPB		20 Jan 2022 11:22:34	U	SMPL
N1187-09MSD - 1	600492261MSD	14487	3.6496	-7470A	PPB		20 Jan 2022 11:24:50	U	SMPL
PB142113TB - 1	PB142113TB	-5	-0.0284	-7470A	PPB		20 Jan 2022 11:27:05	U	SMPL
PB142115TB - 1	PB142115TB	27	-0.0203	-7470A	PPB		20 Jan 2022 11:29:21	U	SMPL
N1176-02LX5 - 1		4	-0.0262	-7470A	PPB		20 Jan 2022 11:31:37	U	SMPL
N1176-02A - 1		15780	3.9777	-7470A	PPB		20 Jan 2022 11:33:53	U	SMPL
N1178-03LX5 - 1		-39	-0.0371	-7470A	PPB		20 Jan 2022 11:36:09	U	SMPL
N1178-03A - 1		18139	4.5764	-7470A	PPB		20 Jan 2022 11:38:25	U	SMPL
N1187-09LX5 - 1		-48	-0.0394	-7470A	PPB		20 Jan 2022 11:40:42	U	SMPL
N1187-09A - 1		15858	3.9975	-7470A	PPB		20 Jan 2022 11:42:59	U	SMPL
N1187-08DLX5 - 1		24214	6.1182	-7470A	PPB		20 Jan 2022 11:45:15	U	SMPL
CCV34 - 1	CCV34	20453	5.1637	-7470A	PPB		20 Jan 2022 11:47:31	U	SMPL
CCB34 - 1	CCB34	-22	-0.0328	-7470A	PPB		20 Jan 2022 11:49:46	U	SMPL