



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118302

Slope : 98.4

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/20/2022	11:45
2	CAL2	1	Water	NA	NA	20.3	7.00	01/20/2022	11:46
3	CAL3	1	Water	NA	NA	20.3	10.03	01/20/2022	11:47
4	ICV	1	Water	NA	NA	20.4	6.99	01/20/2022	11:49
5	CCV1	1	Water	NA	NA	20.4	2.01	01/20/2022	11:50
6	N1209-01	1	Water	NA	NA	20.1	5.59	01/20/2022	12:00
7	N1209-01DUP	1	Water	NA	NA	20.2	5.60	01/20/2022	12:02
8	N1209-02	1	Water	NA	NA	20.2	4.86	01/20/2022	12:05
9	N1209-03	1	Water	NA	NA	20.4	5.47	01/20/2022	12:06
10	CCV2	1	Water	NA	NA	20.4	12.02	01/20/2022	12:10

LB 118302

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1204 WorkList ID : 156165 Department : Wet-Chemistry Date : 01-20-2022 11:35:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/19/2022	Water	N1209-01	pH	Cool 4 deg C	STER09	H11	CELL-A-POOL-WATER(D	01/19/2022	9040C
01/19/2022	Water	N1209-02	pH	Cool 4 deg C	STER09	H11	CELL-B-C-POOL-WATER	01/19/2022	9040C
01/19/2022	Water	N1209-03	pH	Cool 4 deg C	STER09	H11	CELL-D-POOL-WATER-(I	01/19/2022	9040C

11:40

JP

Date/Time 01.20.22
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: 12/06/22

12:45

Date/Time 01.20.22
Raw Sample Received by: 12/06/22
Raw Sample Relinquished by: JP (WC)