



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118317

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	01/20/2022	17:00
2	CAL2	1	Water	NA	NA	20.4	7.00	01/20/2022	17:01
3	CAL3	1	Water	NA	NA	20.4	10.03	01/20/2022	17:02
4	ICV	1	Water	NA	NA	20.5	6.99	01/20/2022	17:05
5	CCV1	1	Water	NA	NA	20.5	2.01	01/20/2022	17:06
6	N1213-02	1	Solid	20.03	20	20.6	6.23	01/20/2022	17:10
7	N1213-05	1	Solid	20.05	20	20.4	6.62	01/20/2022	17:12
8	N1218-02	1	Solid	20.02	20	20.4	7.51	01/20/2022	17:15
9	N1224-08	1	Solid	20.03	20	20.3	9.59	01/20/2022	17:19
10	N1224-16	1	Solid	20.02	20	20.6	9.01	01/20/2022	17:22
11	N1224-16DUP	1	Solid	20.03	20	20.7	9.02	01/20/2022	17:23
12	CCV2	1	Water	NA	NA	20.6	12.03	01/20/2022	17:25

WORKLIST(Hardcopy Internal Chain)

118317

WorkList Name : CORROSIVITY N1218 WorkList ID : 156186 Department : Wet-Chemistry Date : 01-20-2022 17:07:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/20/2022	Solid	N1213-02	Corrosivity	Cool 4 deg C	PSEG03	D11	CHRT-27080	01/20/2022	9045D
01/20/2022	Solid	N1213-05	Corrosivity	Cool 4 deg C	PSEG03	D11	OILY-STONES-DRUM	01/20/2022	9045D
01/20/2022	Solid	N1218-02	Corrosivity	Cool 4 deg C	ENV127	H11	EAST-BOILER-PIT-WAS1	01/20/2022	9045D
01/20/2022	Solid	N1224-08	Corrosivity	Cool 4 deg C	ENV127	D11	AREA-2-COMPOSITE	01/20/2022	9045D
01/20/2022	Solid	N1224-16	Corrosivity	Cool 4 deg C	ENV127	D11	AREA-1-COMPOSITE	01/20/2022	9045D

Date/Time 01.20.22 16:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 01.20.22 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]