


7470A

LB118312
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38727



A= 0.0000e+000

B= 2.6195e-004

C= -1.0151e-001

Rho= 0.9997626

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	-0.045	-0.045	214	0.000	214					-13
0.2	0.200	0.174	-0.026	1053	0.0 %	1053					4
2.5	2.500	2.602	0.102	10322	0.0 %	10322					1
5.0	5.000	5.065	0.065	19724	0.0 %	19724					-2
7.5	7.500	7.360	-0.140	28486	0.0 %	28486					0
10.0	10.000	10.043	0.043	38727	0.0 %	38727					

LB118318 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jan 2022 11:04:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	214	-	0.0000	7470A	PPB	21 Jan 2022 11:30:16	S	Std
0.2 - 1	50.2	1053	-	0.2000	7470A	PPB	21 Jan 2022 11:32:33	S	Std
2.5 - 1	52.5	10322	-	2.5000	7470A	PPB	21 Jan 2022 11:37:00	S	Std
5.0 - 1	55	19724	-	5.0000	7470A	PPB	21 Jan 2022 11:39:17	S	Std
7.5 - 1	57.5	28486	-	7.5000	7470A	PPB	21 Jan 2022 11:41:33	S	Std
10.0 - 1	510	38727	-	10.0000	7470A	PPB	21 Jan 2022 11:47:14	S	Std
ICV53 - 1	ICV53	14973	3.8207	-	7470A	PPB	21 Jan 2022 11:53:07	U	SMPL
ICB53 - 1	ICB53	104	-0.0743	-	7470A	PPB	21 Jan 2022 11:55:23	U	SMPL
CCV40 - 1	CCV40	19415	4.9843	-	7470A	PPB	21 Jan 2022 11:57:37	U	SMPL
CCB40 - 1	CCB40	100	-0.0753	-	7470A	PPB	21 Jan 2022 11:59:53	U	SMPL
CRA - 1	CRA	1051	0.1738	-	7470A	PPB	21 Jan 2022 12:02:07	U	SMPL
HighStd - 1	HighStd	39105	10.1421	-	7470A	PPB	21 Jan 2022 12:04:22	U	SMPL
ChkStd - 1	ChkStd	26595	6.8651	-	7470A	PPB	21 Jan 2022 12:06:37	U	SMPL
PB142168BL - 1	PBW	15	-0.0976	-	7470A	PPB	21 Jan 2022 12:11:26	U	SMPL
PB142168BS - 1	LCSW	15680	4.0059	-	7470A	PPB	21 Jan 2022 12:13:42	U	SMPL
N1186-08 - 1	17103	2369	0.5191	-	7470A	PPB	21 Jan 2022 12:15:58	U	SMPL
N1188-02 - 1	72-11940	717	0.0863	-	7470A	PPB	21 Jan 2022 12:23:18	U	SMPL
N1188-02DUP - 1	72-11940DUP	621	0.0612	-	7470A	PPB	21 Jan 2022 12:25:35	U	SMPL
N1188-02MS - 1	72-11940MS	15148	3.8665	-	7470A	PPB	21 Jan 2022 12:27:52	U	SMPL
N1188-02MSD - 1	72-11940MSD	15437	3.9422	-	7470A	PPB	21 Jan 2022 12:32:46	U	SMPL
CCV41 - 1	CCV41	21366	5.4954	-	7470A	PPB	21 Jan 2022 12:35:01	U	SMPL
CCB41 - 1	CCB41	87	-0.0787	-	7470A	PPB	21 Jan 2022 12:37:16	U	SMPL
N1188-02LX5 - 1		278	-0.0287	-	7470A	PPB	21 Jan 2022 12:39:32	U	SMPL
N1188-02A - 1		15916	4.0677	-	7470A	PPB	21 Jan 2022 12:41:47	U	SMPL
PB142136TB - 1	PB142136TB	72	-0.0826	-	7470A	PPB	21 Jan 2022 12:44:02	U	SMPL
CCV42 - 1	CCV42	20649	5.3075	-	7470A	PPB	21 Jan 2022 12:46:18	U	SMPL
CCB42 - 1	CCB42	103	-0.0745	-	7470A	PPB	21 Jan 2022 12:48:34	U	SMPL