

LB118326


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37540



A= 0.0000e+000

B= 2.6350e-004

C= -1.0302e-001

Rho= 0.9992712

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.026	-0.026	291	0.000	291					-13
0.2	0.200	0.175	-0.025	1056	0.0 %	1056					0
2.5	2.500	2.495	-0.005	9861	0.0 %	9861					0
5.0	5.000	4.997	-0.003	19355	0.0 %	19355					4
7.5	7.500	7.770	0.270	29880	0.0 %	29880					-2
10.0	10.000	9.789	-0.211	37540	0.0 %	37540					

LB118326 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jan 2022 14:51:28

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	291	-	0.0007470A	PPB	21 Jan 2022 14:52:11	S	Std
0.2 - 1	51.2	1056	-	0.2007470A	PPB	21 Jan 2022 14:57:41	S	Std
2.5 - 1	52.5	9861	-	2.5007470A	PPB	21 Jan 2022 14:59:58	S	Std
5.0 - 1	55	19355	-	5.0007470A	PPB	21 Jan 2022 15:02:15	S	Std
7.5 - 1	57.5	29880	-	7.5007470A	PPB	21 Jan 2022 15:04:31	S	Std
10.0 - 1	58	37540	-	10.0007470A	PPB	21 Jan 2022 15:06:47	S	Std
ICV54 - 1	ICV54	14936	3.8326	-7470A	PPB	21 Jan 2022 15:09:30	U	SMPL
ICB54 - 1	ICB54	395	0.0011	-7470A	PPB	21 Jan 2022 15:19:33	U	SMPL
CCV43 - 1	CCV43	17979	4.6344	-7470A	PPB	21 Jan 2022 15:21:48	U	SMPL
CCB43 - 1	CCB43	213	-0.0469	-7470A	PPB	21 Jan 2022 15:24:03	U	SMPL
CRA - 1	CRA	1034	0.1694	-7470A	PPB	21 Jan 2022 15:26:18	U	SMPL
HighStd - 1	HighStd	38971	10.1657	-7470A	PPB	21 Jan 2022 15:28:33	U	SMPL
ChkStd - 1	ChkStd	29059	7.5539	-7470A	PPB	21 Jan 2022 15:30:48	U	SMPL
PB142181BL - 1	PBW	302	-0.0234	-7470A	PPB	21 Jan 2022 15:33:04	U	SMPL
PB142181BS - 1	LCSW	13546	3.4663	-7470A	PPB	21 Jan 2022 15:35:20	U	SMPL
N1215-03 - 1	31460	396	0.0013	-7470A	PPB	21 Jan 2022 15:37:36	U	SMPL
N1218-02 - 1	EAST-BOILER-PIT-WASTE-CLASS	354	-0.0097	-7470A	PPB	21 Jan 2022 15:39:52	U	SMPL
N1224-08 - 1	AREA-2-COMPOSITE	342	-0.0129	-7470A	PPB	21 Jan 2022 15:42:10	U	SMPL
N1224-16 - 1	AREA-1-COMPOSITE	376	-0.0039	-7470A	PPB	21 Jan 2022 15:44:26	U	SMPL
N1224-16DUP - 1	AREA-1-COMPOSITEDUP	370	-0.0055	-7470A	PPB	21 Jan 2022 15:49:47	U	SMPL
CCV44 - 1	CCV44	19899	5.1403	-7470A	PPB	21 Jan 2022 15:52:02	U	SMPL
CCB44 - 1	CCB44	358	-0.0087	-7470A	PPB	21 Jan 2022 15:56:48	U	SMPL
N1224-16MS - 1	AREA-1-COMPOSITEMS	14242	3.6497	-7470A	PPB	21 Jan 2022 16:01:38	U	SMPL
N1224-16MSD - 1	AREA-1-COMPOSITMSD	14616	3.7482	-7470A	PPB	21 Jan 2022 16:03:53	U	SMPL
PB142182BL - 1	PBW	211	-0.0474	-7470A	PPB	21 Jan 2022 16:06:08	U	SMPL
PB142182BS - 1	LCSW	15358	3.9438	-7470A	PPB	21 Jan 2022 16:10:58	U	SMPL
N1210-02 - 1	20220030-KEANSBURG-1-MODIFIED-ELUTRIATE	640	0.0656	-7470A	PPB	21 Jan 2022 16:13:13	U	SMPL
N1210-02DUP - 1	20220030-KEANSBURG-1-MODIFIED-ELUTRIATEDUP	742	0.0925	-7470A	PPB	21 Jan 2022 16:15:30	U	SMPL
N1210-02MS - 1	20220030-KEANSBURG-1-MODIFIED-ELUTRIATEMS	16122	4.1451	-7470A	PPB	21 Jan 2022 16:17:46	U	SMPL
N1210-02MSD - 1	20220030-KEANSBURG-1-MODIFIED-ELUTRIATEMSD	16451	4.2318	-7470A	PPB	21 Jan 2022 16:20:03	U	SMPL
N1210-08 - 1	20220030-KEANSBURG-1-MODIFIED-ELUTRIATE	1026	0.1673	-7470A	PPB	21 Jan 2022 16:22:20	U	SMPL
PB142195BL - 1	PBW	239	-0.0400	-7470A	PPB	21 Jan 2022 16:27:19	U	SMPL
CCV45 - 1	CCV45	18723	4.8304	-7470A	PPB	21 Jan 2022 16:29:34	U	SMPL
CCB45 - 1	CCB45	315	-0.0200	-7470A	PPB	21 Jan 2022 16:31:49	U	SMPL
PB142195BS - 1	LCSW	15600	4.0075	-7470A	PPB	21 Jan 2022 16:36:49	U	SMPL
N1217-02 - 1	2250	491	0.0264	-7470A	PPB	21 Jan 2022 16:39:04	U	SMPL
N1217-05 - 1	2256-2261	376	-0.0039	-7470A	PPB	21 Jan 2022 16:41:20	U	SMPL
N1217-05DUP - 1	2256-2261DUP	263	-0.0337	-7470A	PPB	21 Jan 2022 16:43:35	U	SMPL
N1217-05MS - 1	2256-2261MS	15522	3.9870	-7470A	PPB	21 Jan 2022 16:52:34	U	SMPL
N1217-05MSD - 1	2256-2261MSD	15847	4.0726	-7470A	PPB	21 Jan 2022 16:54:51	U	SMPL
PB142169TB - 1	PB142169TB	165	-0.0595	-7470A	PPB	21 Jan 2022 16:59:44	U	SMPL
PB142193TB - 1	PB142193TB	219	-0.0453	-7470A	PPB	21 Jan 2022 17:02:00	U	SMPL
CCV46 - 1	CCV46	18382	4.7406	-7470A	PPB	21 Jan 2022 17:04:17	U	SMPL
CCB46 - 1	CCB46	476	0.0224	-7470A	PPB	21 Jan 2022 17:06:33	U	SMPL
N1224-16LX5 - 1		128	-0.0693	-7470A	PPB	21 Jan 2022 17:08:49	U	SMPL
N1224-16A - 1		16027	4.1200	-7470A	PPB	21 Jan 2022 17:11:04	U	SMPL
N1210-02LX5 - 1		1056	0.1752	-7470A	PPB	21 Jan 2022 17:13:19	U	SMPL
N1210-02A - 1		16621	4.2765	-7470A	PPB	21 Jan 2022 17:15:35	U	SMPL
N1217-05LX5 - 1		40	-0.0925	-7470A	PPB	21 Jan 2022 17:17:50	U	SMPL
N1217-05A - 1		16563	4.2613	-7470A	PPB	21 Jan 2022 17:20:06	U	SMPL
CCV47 - 1	CCV47	19670	5.0793	-7470A	PPB	21 Jan 2022 17:22:21	U	SMPL
CCB47 - 1	CCB47	290	-0.0266	-7470A	PPB	21 Jan 2022 17:24:38	U	SMPL