



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 01/24/2022OVEN IN TEMP1: 103 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 01/21/2022 17:00 DATE & TIME OUT1: 01/24/2022 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 01/24/2022 08:30 DATE & TIME OUT2: 01/24/2022 10:00Run Number: LB118331BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB118331BL	LB118331BL	78.6369	78.6369	100	78.6369	78.6369	78.6369	0.0000	0
2	N1228-03	1245-N43502	81.4748	81.4748	100	81.7826	81.7826	81.7826	0.3078	3078
3	N1228-03DUP	1245-N43502DUP	75.5431	75.5431	100	75.8538	75.8538	75.8538	0.3107	3107

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

171833A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS WN1236

Department : Wet-Chemistry

WorkList ID : 156224

Date : 01-21-2022 16:36:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/26/2022	Water	N1228-03	TS	Cool 4 deg C	PSEG03	D12	1245-N43502	01/21/2022	SM2540 B

Date/Time 01.21.22 16:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 01.21.22 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]