



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118333

Slope : 98.7

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/21/2022	16:47
2	CAL2	1	Water	NA	NA	20.6	7.01	01/21/2022	16:48
3	CAL3	1	Water	NA	NA	20.6	10.03	01/21/2022	16:49
4	ICV	1	Water	NA	NA	20.6	6.98	01/21/2022	16:50
5	CCV1	1	Water	NA	NA	20.6	2.01	01/21/2022	16:51
6	N1228-01	1	Water	NA	NA	20.2	7.81	01/21/2022	17:00
7	N1228-02	1	Water	NA	NA	20.3	7.64	01/21/2022	17:01
8	N1228-03	1	Water	NA	NA	20.5	7.73	01/21/2022	17:03
9	N1228-03DUP	1	Water	NA	NA	20.6	7.74	01/21/2022	17:04
10	CCV2	1	Water	NA	NA	20.5	12.03	01/21/2022	17:10

W118333

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1236 WorkList ID : 156223 Department : Wet-Chemistry Date : 01-21-2022 16:36:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/21/2022	Water	N1228-01	pH	Cool 4 deg C	PSEG03	D12	1245	01/21/2022	9040C
01/21/2022	Water	N1228-02	pH	Cool 4 deg C	PSEG03	D12	N43502	01/21/2022	9040C
01/21/2022	Water	N1228-03	pH	Cool 4 deg C	PSEG03	D12	1245-N43502	01/21/2022	9040C

Date/Time 01.21.22 16:45
 Raw Sample Received by: JN (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 01.21.22 18:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JN (WC)