



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118357

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	01/24/2022	13:58
2	N1228-03		1	12	100.00	>212.0	>212.0	01/24/2022	14:25
3	N1228-03DUP		1	12	100.00	>212.0	>212.0	01/24/2022	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB118357

WorkList Name : fp12422-1 WorkList ID : 156255 Department : Wet-Chemistry Date : 01-24-2022 09:25:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/26/2022	Water	N1228-03	Flash Point	Cool 4 deg C	PSEG03	D12	1245-N43502	01/21/2022	1010B

Date/Time 01/24/22 13:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 01/24/22 14:50
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)