

LB118360


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43459



A= 0.0000e+000

B= 2.3213e-004

C= 1.6438e-003

Rho= 0.9998539

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	0.016	0.016	60	0.000	60					-
0.2	0.200	0.205	0.005	877	0.0 %	877					-2
2.5	2.500	2.546	0.046	10960	0.0 %	10960					2
5.0	5.000	4.946	-0.054	21299	0.0 %	21299					-1
7.5	7.500	7.398	-0.102	31861	0.0 %	31861					-1
10.0	10.000	10.090	0.090	43459	0.0 %	43459					1

LB118360 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jan 2022 09:23:03

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	60	-	0.0007470A	PPB		25 Jan 2022 09:27:22	S	Std
0.2 - 1	50.2	877	-	0.20007470A	PPB		25 Jan 2022 09:29:39	S	Std
2.5 - 1	52.5	10960	-	2.50007470A	PPB		25 Jan 2022 09:31:56	S	Std
5.0 - 1	55	21299	-	5.00007470A	PPB		25 Jan 2022 09:34:13	S	Std
7.5 - 1	57.5	31861	-	7.50007470A	PPB		25 Jan 2022 09:36:30	S	Std
10.0 - 1	59.5	43459	-	10.00007470A	PPB		25 Jan 2022 09:41:18	S	Std
ICV58 - 1	ICV58	17479	4.0591	-7470A	PPB		25 Jan 2022 09:46:31	U	SMPL
ICB58 - 1	ICB58	-21	-0.0032	-7470A	PPB		25 Jan 2022 09:48:46	U	SMPL
CCV58 - 1	CCV58	19382	4.5008	-7470A	PPB		25 Jan 2022 09:51:01	U	SMPL
CCB58 - 1	CCB58	-27	-0.0046	-7470A	PPB		25 Jan 2022 09:53:16	U	SMPL
CRA - 1	CRA	846	0.1980	-7470A	PPB		25 Jan 2022 09:55:30	U	SMPL
HighStd - 1	HighStd	43010	9.9857	-7470A	PPB		25 Jan 2022 09:57:46	U	SMPL
ChkStd - 1	ChkStd	28515	6.6209	-7470A	PPB		25 Jan 2022 10:00:01	U	SMPL
PB142233BL - 1	PBW	-52	-0.0104	-7470A	PPB		25 Jan 2022 10:02:18	U	SMPL
N1210-06 - 1	20220027-KEANSBURG-1-COMP-SPLP-LEACHATE	131	0.0321	-7470A	PPB		25 Jan 2022 10:06:51	U	SMPL
N1210-06DUP - 1	20220027-KEANSBURG-1-COMP-SPLP-LEACHATEDUP	115	0.0283	-7470A	PPB		25 Jan 2022 10:09:07	U	SMPL
N1210-06MS - 1	20220027-KEANSBURG-1-COMP-SPLP-LEACHATEMS	15396	3.5756	-7470A	PPB		25 Jan 2022 10:14:02	U	SMPL
N1210-06MSD - 1	20220027-KEANSBURG-1-COMP-SPLP-LEACHATEMSD	15395	3.5753	-7470A	PPB		25 Jan 2022 10:16:19	U	SMPL
PB142233BS - 1	LCSW	15749	3.6575	-7470A	PPB		25 Jan 2022 10:18:51	U	SMPL
PB142234BL - 1	PBW	-34	-0.0062	-7470A	PPB		25 Jan 2022 10:21:23	U	SMPL
CCV59 - 1	CCV59	20661	4.7977	-7470A	PPB		25 Jan 2022 10:23:38	U	SMPL
CCB59 - 1	CCB59	-23	-0.0037	-7470A	PPB		25 Jan 2022 10:25:53	U	SMPL
PB142234BS - 1	LCSW	16009	3.7179	-7470A	PPB		25 Jan 2022 10:30:39	U	SMPL
N1235-02 - 1	TP-1	-41	-0.0079	-7470A	PPB		25 Jan 2022 10:32:54	U	SMPL
N1235-05 - 1	TP-2	17	0.0056	-7470A	PPB		25 Jan 2022 10:35:09	U	SMPL
N1242-02 - 1	PSC-124042	28	0.0081	-7470A	PPB		25 Jan 2022 10:37:25	U	SMPL
N1242-04 - 1	PSC-124055	34	0.0095	-7470A	PPB		25 Jan 2022 10:39:40	U	SMPL
N1242-06 - 1	PSC-124035	26	0.0077	-7470A	PPB		25 Jan 2022 10:41:57	U	SMPL
N1243-03 - 1	AMI-1	34	0.0095	-7470A	PPB		25 Jan 2022 10:44:14	U	SMPL
N1243-07 - 1	AMI-2	34	0.0095	-7470A	PPB		25 Jan 2022 10:46:30	U	SMPL
N1243-11 - 1	AMI-3	28	0.0081	-7470A	PPB		25 Jan 2022 10:48:47	U	SMPL
N1243-11DUP - 1	AMI-3DUP	32	0.0091	-7470A	PPB		25 Jan 2022 10:51:03	U	SMPL
CCV60 - 1	CCV60	20437	4.7457	-7470A	PPB		25 Jan 2022 10:53:18	U	SMPL
CCB60 - 1	CCB60	-4	0.0007	-7470A	PPB		25 Jan 2022 10:55:34	U	SMPL
N1243-11MS - 1	AMI-3MS	17052	3.9600	-7470A	PPB		25 Jan 2022 10:57:50	U	SMPL
N1243-11MSD - 1	AMI-3MSD	17104	3.9720	-7470A	PPB		25 Jan 2022 11:00:05	U	SMPL
PB142235BL - 1	PBW	-30	-0.0053	-7470A	PPB		25 Jan 2022 11:02:20	U	SMPL
PB142235BS - 1	LCSW	17111	3.9737	-7470A	PPB		25 Jan 2022 11:04:36	U	SMPL
N1228-04 - 1	1245-N43502	-26	-0.0044	-7470A	PPB		25 Jan 2022 11:06:52	U	SMPL
N1228-04DUP - 1	1245-N43502DUP	24	0.0072	-7470A	PPB		25 Jan 2022 11:09:08	U	SMPL
N1228-04MS - 1	1245-N43502MS	17131	3.9783	-7470A	PPB		25 Jan 2022 11:19:25	U	SMPL
N1228-04MSD - 1	1245-N43502MSD	17110	3.9734	-7470A	PPB		25 Jan 2022 11:21:42	U	SMPL
PB142236BL - 1	PBW	-22	-0.0035	-7470A	PPB		25 Jan 2022 11:24:00	U	SMPL
PB142236BS - 1	LCSW	17153	3.9834	-7470A	PPB		25 Jan 2022 11:26:15	U	SMPL
CCV61 - 1	CCV61	20674	4.8008	-7470A	PPB		25 Jan 2022 11:28:32	U	SMPL
CCB61 - 1	CCB61	-24	-0.0039	-7470A	PPB		25 Jan 2022 11:30:47	U	SMPL
N1228-03 - 1	1245-N43502	3458	0.8044	-7470A	PPB		25 Jan 2022 11:33:03	U	SMPL
N1228-03DUP - 1	1245-N43502DUP	3322	0.7728	-7470A	PPB		25 Jan 2022 11:35:18	U	SMPL
N1228-03MS - 1	1245-N43502MS	18574	4.3133	-7470A	PPB		25 Jan 2022 11:39:49	U	SMPL
N1228-03MSD - 1	1245-N43502MSD	17985	4.1766	-7470A	PPB		25 Jan 2022 11:42:04	U	SMPL
PB142202TB - 1	PB142202TB	-3	0.0009	-7470A	PPB		25 Jan 2022 11:44:21	U	SMPL
PB142208TB - 1	PB142208TB	17	0.0056	-7470A	PPB		25 Jan 2022 11:46:37	U	SMPL
PB142223TB - 1	PB142223TB	35	0.0098	-7470A	PPB		25 Jan 2022 11:48:53	U	SMPL
N1210-06LX5 - 1		187	0.0451	-7470A	PPB		25 Jan 2022 11:51:10	U	SMPL
N1210-06A - 1		17173	3.9881	-7470A	PPB		25 Jan 2022 11:53:27	U	SMPL
CCV62 - 1	CCV62	20587	4.7806	-7470A	PPB		25 Jan 2022 11:55:44	U	SMPL
CCB62 - 1	CCB62	-9	-0.0004	-7470A	PPB		25 Jan 2022 11:57:59	U	SMPL
N1243-11LX5 - 1		2	0.0021	-7470A	PPB		25 Jan 2022 12:00:15	U	SMPL
N1243-11A - 1		17356	4.0305	-7470A	PPB		25 Jan 2022 12:02:31	U	SMPL
N1228-04LX5 - 1		-32	-0.0058	-7470A	PPB		25 Jan 2022 12:04:46	U	SMPL
N1228-04A - 1		17194	3.9929	-7470A	PPB		25 Jan 2022 12:07:02	U	SMPL
N1228-03LX5 - 1		576	0.1354	-7470A	PPB		25 Jan 2022 12:09:17	U	SMPL
N1228-03A - 1		18327	4.2559	-7470A	PPB		25 Jan 2022 12:11:33	U	SMPL
CCV63 - 1	CCV63	19634	4.5593	-7470A	PPB		25 Jan 2022 12:13:49	U	SMPL
CCB63 - 1	CCB63	-19	-0.0028	-7470A	PPB		25 Jan 2022 12:16:06	U	SMPL