

LB118366


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

57028



A= 0.0000e+000

B= 1.7621e-004

C= 4.9116e-002

Rho= 0.9998355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.057	0.057	46	0.000	46					1
0.2	0.200	0.237	0.037	1067	0.0 %	1067					19
2.5	2.500	2.448	-0.052	13611	0.0 %	13611					2
5.0	5.000	4.934	-0.066	27720	0.0 %	27720					1
7.5	7.500	7.426	-0.074	41865	0.0 %	41865					1
10.0	10.000	10.098	0.098	57028	0.0 %	57028					1

LB118366 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Jan 2022 14:19:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	46	-	0.00007471B	PPB		25 Jan 2022 14:26:30	S	Std
0.2 - 1	50.2	1067	-	0.20007471B	PPB		25 Jan 2022 14:28:47	S	Std
2.5 - 1	52.5	13611	-	2.50007471B	PPB		25 Jan 2022 14:31:04	S	Std
5.0 - 1	55	27720	-	5.00007471B	PPB		25 Jan 2022 14:33:21	S	Std
7.5 - 1	57.5	41865	-	7.50007471B	PPB		25 Jan 2022 14:35:37	S	Std
10.0 - 1	510	57028	-	10.00007471B	PPB		25 Jan 2022 14:43:03	S	Std
ICV59 - 1	ICV59	21741	3.8801	-7471B	PPB		25 Jan 2022 14:46:04	U	SMPL
ICB59 - 1	ICB59	-31	0.0437	-7471B	PPB		25 Jan 2022 14:48:19	U	SMPL
CCV64 - 1	CCV64	26760	4.7646	-7471B	PPB		25 Jan 2022 14:50:35	U	SMPL
CCB64 - 1	CCB64	-51	0.0401	-7471B	PPB		25 Jan 2022 14:52:50	U	SMPL
CRA - 1	CRA	999	0.2252	-7471B	PPB		25 Jan 2022 14:55:05	U	SMPL
HighStd - 1	HighStd	57988	10.2673	-7471B	PPB		25 Jan 2022 14:57:20	U	SMPL
ChkStd - 1	ChkStd	37454	6.6490	-7471B	PPB		25 Jan 2022 14:59:35	U	SMPL
PB142270BL - 1	PBS	-79	0.0352	-7471B	PPB		25 Jan 2022 15:01:53	U	SMPL
PB142270BS - 1	LCSS	20075	3.5866	-7471B	PPB		25 Jan 2022 15:04:09	U	SMPL
N1245-01 - 1	SB-1(0.5-1)	31477	5.5957	-7471B	PPB		25 Jan 2022 15:06:26	U	SMPL
N1245-01DUP - 1	SB-1(0.5-1)DUP	30349	5.3970	-7471B	PPB		25 Jan 2022 15:08:42	U	SMPL
N1245-01MS - 1	SB-1(0.5-1)MS	53512	9.4786	-7471B	PPB		25 Jan 2022 15:13:54	U	SMPL
N1245-01MSD - 1	SB-1(0.5-1)MSD	54975	9.7364	-7471B	PPB		25 Jan 2022 15:16:10	U	SMPL
N1245-02 - 1	SB-1(4-4.5)	3561	0.6766	-7471B	PPB		25 Jan 2022 15:18:27	U	SMPL
CCV65 - 1	CCV65	26107	4.6495	-7471B	PPB		25 Jan 2022 15:20:46	U	SMPL
CCB65 - 1	CCB65	-35	0.0429	-7471B	PPB		25 Jan 2022 15:23:01	U	SMPL
N1245-03 - 1	SB-2(0.5-1)	17617	3.1534	-7471B	PPB		25 Jan 2022 15:25:17	U	SMPL
N1245-04 - 1	SB-2(2.5-3)	10535	1.9055	-7471B	PPB		25 Jan 2022 15:27:32	U	SMPL
N1245-05 - 1	SB-3(0.5-1)	26666	4.7480	-7471B	PPB		25 Jan 2022 15:29:47	U	SMPL
N1245-06 - 1	SB-3(3-3.5)	17818	3.1889	-7471B	PPB		25 Jan 2022 15:32:03	U	SMPL
N1245-07 - 1	SB-4(0.5-1)	47005	8.3320	-7471B	PPB		25 Jan 2022 15:34:18	U	SMPL
N1245-08 - 1	SB-4(1.5-2)	24210	4.3152	-7471B	PPB		25 Jan 2022 15:36:34	U	SMPL
N1245-09 - 1	SB-5(0.5-1)	22811	4.0687	-7471B	PPB		25 Jan 2022 15:38:52	U	SMPL
N1245-10 - 1	SB-5(3.5-4)	11848	2.1369	-7471B	PPB		25 Jan 2022 15:41:09	U	SMPL
N1247-01 - 1	SS-1	1946	0.3920	-7471B	PPB		25 Jan 2022 15:43:25	U	SMPL
N1248-01 - 1	C1-C2	527251	HIGH	-7471B	PPB		25 Jan 2022 15:45:41	X	
CCV66 - 1	CCV66	25876	4.6088	-7471B	PPB		25 Jan 2022 15:47:29	U	SMPL
CCB66 - 1	CCB66	-89	0.0334	-7471B	PPB		25 Jan 2022 15:49:48	U	SMPL
N1251-01 - 1	VNJ-222	932	0.2133	-7471B	PPB		25 Jan 2022 15:52:04	U	SMPL
N1251-03 - 1	275	1157	0.2530	-7471B	PPB		25 Jan 2022 15:54:19	U	SMPL
N1248-01DLX10 - 1		90547	16.0046	-7471B	PPB		25 Jan 2022 15:56:35	U	SMPL
N1245-01LX5 - 1		5775	1.0667	-7471B	PPB		25 Jan 2022 15:58:50	U	SMPL
N1245-01A - 1		55833	9.8876	-7471B	PPB		25 Jan 2022 16:01:09	U	SMPL
N1248-01DL2X50 - 1		18848	3.3704	-7471B	PPB		25 Jan 2022 16:03:25	U	SMPL
CCV67 - 1	CCV67	26968	4.8012	-7471B	PPB		25 Jan 2022 16:05:43	U	SMPL
CCB67 - 1	CCB67	15	0.0518	-7471B	PPB		25 Jan 2022 16:10:37	U	SMPL