



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118368

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/25/2022	14:00
2	CAL2	1	Water	NA	NA	20.3	7.00	01/25/2022	14:01
3	CAL3	1	Water	NA	NA	20.3	10.03	01/25/2022	14:02
4	ICV	1	Water	NA	NA	20.4	6.99	01/25/2022	14:05
5	CCV1	1	Water	NA	NA	20.4	2.01	01/25/2022	14:06
6	N1253-01	1	Water	NA	NA	20.6	6.14	01/25/2022	14:10
7	N1253-01DUP	1	Water	NA	NA	20.7	6.15	01/25/2022	14:11
8	N1257-01	1	Water	NA	NA	20.5	6.52	01/25/2022	14:12
9	N1257-02	1	Water	NA	NA	20.2	6.79	01/25/2022	14:15
10	CCV2	1	Water	NA	NA	20.4	12.03	01/25/2022	14:19

VB 118368

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1257 WorkList ID : 156295 Department : Wet-Chemistry Date : 01-25-2022 13:47:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/24/2022	Water	N1253-01	A pH	Cool 4 deg C	PSEG03	D12	NWB-1732	01/24/2022	9040C
01/25/2022	Water	N1257-01	pH	Cool 4 deg C	PSEG03	L11	CPIT-1	01/25/2022	9040C
01/25/2022	Water	N1257-02	pH	Cool 4 deg C	PSEG03	L11	CPIT-2	01/25/2022	9040C

Date/Time 01-25-22 13:50
Raw Sample Received by: 70 (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 01-25-22 15:20
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: 70 (WC)