



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118378

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/26/2022	11:25
2	CAL2	1	Water	NA	NA	20.7	7.00	01/26/2022	11:26
3	CAL3	1	Water	NA	NA	20.7	10.03	01/26/2022	11:27
4	ICV	1	Water	NA	NA	20.8	7.02	01/26/2022	11:29
5	CCV1	1	Water	NA	NA	20.7	2.01	01/26/2022	11:30
6	N1274-01	1	Water	NA	NA	20.5	6.62	01/26/2022	11:33
7	N1274-01DUP	1	Water	NA	NA	20.6	6.63	01/26/2022	11:34
8	CCV2	1	Water	NA	NA	20.7	12.03	01/26/2022	11:35

VB 118378

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1274 WorkList ID : 156318 Department : Wet-Chemistry Date : 01-26-2022 11:13:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2022	Water	N1274-01	pH	Cool 4 deg C	STER10	D11	CELL-A-POOL-WATER(D	01/10/2022	9040C

Date/Time 01-26-2022 11:20
Raw Sample Received by: JD (CC)
Raw Sample Relinquished by: JD (CC)

Date/Time 01-26-22 12:30
Raw Sample Received by: JD (CC)
Raw Sample Relinquished by: JD (CC)