



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB118380

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP92563

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB118380BL		1	21.3	0.84	0.90	01/26/2022	12:52
2	LB118380BS	1413	1	21.5	1315.00	1409.21	01/26/2022	12:53
3	N1274-01		1	17.4	1.50	1.75	01/26/2022	12:55
4	N1274-01DUP		1	17.5	1.51	1.76	01/26/2022	12:56

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

WORKLIST(Hardcopy Internal Chain)

LB 118380

WorkList Name : CONDDUCTANCE-0126; WorkList ID : 156317 Department : Wet-Chemistry Date : 01-26-2022 11:07:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2022	Water	N1274-01	Conductance	Cool 4 deg C	STER10	D11	CELL-A-POOL-WATER(D	01/10/2022	SM2510 B

Date/Time 01/26/22 11:20
Raw Sample Received by: R (WC)
Raw Sample Relinquished by: SA (WC)

Date/Time 01/26/22 12:30
Raw Sample Received by: SA (WC)
Raw Sample Relinquished by: R (WC)