

LB118384


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40365



A= 0.0000e+000

B= 2.5001e-004

C= 1.3526e-003

Rho= 0.9997997

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	46	0.000	46					-8
0.2	0.200	0.184	-0.016	732	0.0 %	732					3
2.5	2.500	2.579	0.079	10309	0.0 %	10309					-1
5.0	5.000	4.956	-0.044	19817	0.0 %	19817					-2
7.5	7.500	7.375	-0.125	29495	0.0 %	29495					1
10.0	10.000	10.093	0.093	40365	0.0 %	40365					

LB118384 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jan 2022 14:04:09

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	46	-	0.00007470A	PPB	26 Jan 2022 14:05:36	S	Std
0.2 - 1	50.2	732	-	0.20007470A	PPB	26 Jan 2022 14:07:53	S	Std
2.5 - 1	52.5	10309	-	2.50007470A	PPB	26 Jan 2022 14:10:10	S	Std
5.0 - 1	55	19817	-	5.00007470A	PPB	26 Jan 2022 14:12:27	S	Std
7.5 - 1	57.5	29495	-	7.50007470A	PPB	26 Jan 2022 14:14:43	S	Std
10.0 - 1	510	40365	-	10.00007470A	PPB	26 Jan 2022 14:24:37	S	Std
ICV61 - 1	ICV61	17036	4.2605	-7470A	PPB	26 Jan 2022 14:32:17	U	SMPL
ICB61 - 1	ICB61	-24	-0.0046	-7470A	PPB	26 Jan 2022 14:34:32	U	SMPL
CCV72 - 1	CCV72	19177	4.7958	-7470A	PPB	26 Jan 2022 14:36:47	U	SMPL
CCB72 - 1	CCB72	-37	-0.0079	-7470A	PPB	26 Jan 2022 14:39:03	U	SMPL
CRA - 1	CRA	710	0.1789	-7470A	PPB	26 Jan 2022 14:41:17	U	SMPL
HighStd - 1	HighStd	41115	10.2805	-7470A	PPB	26 Jan 2022 14:43:33	U	SMPL
ChkStd - 1	ChkStd	26723	6.6823	-7470A	PPB	26 Jan 2022 14:45:48	U	SMPL
PB142290BL - 1	PBW	-57	-0.0129	-7470A	PPB	26 Jan 2022 14:48:05	U	SMPL
PB142290BS - 1	LCSW	14861	3.7167	-7470A	PPB	26 Jan 2022 14:50:21	U	SMPL
N1248-02 - 1	C1-C2	75	0.0201	-7470A	PPB	26 Jan 2022 14:52:37	U	SMPL
N1263-02 - 1	D2-E2B	53	0.0146	-7470A	PPB	26 Jan 2022 14:54:53	U	SMPL
N1267-02 - 1	RED-EXT-WALL	54	0.0149	-7470A	PPB	26 Jan 2022 14:57:10	U	SMPL
N1267-05 - 1	GRAY-EXT-WALL	69	0.0186	-7470A	PPB	26 Jan 2022 14:59:27	U	SMPL
N1267-08 - 1	RED-BRICK-EXT-WALL	25	0.0076	-7470A	PPB	26 Jan 2022 15:01:42	U	SMPL
CCV73 - 1	CCV73	18943	4.7373	-7470A	PPB	26 Jan 2022 15:03:57	U	SMPL
CCB73 - 1	CCB73	-32	-0.0066	-7470A	PPB	26 Jan 2022 15:06:13	U	SMPL
N1267-11 - 1	ART-ROOM-FLOORS	18	0.0059	-7470A	PPB	26 Jan 2022 15:08:28	U	SMPL
N1267-14 - 1	BASMENT-FLOORS	26	0.0079	-7470A	PPB	26 Jan 2022 15:10:44	U	SMPL
N1267-17 - 1	BASEMENT-WALLS	29	0.0086	-7470A	PPB	26 Jan 2022 15:13:00	U	SMPL
N1267-20 - 1	2ND-3RD-FLOOR-WAKKS-1	28	0.0084	-7470A	PPB	26 Jan 2022 15:15:16	U	SMPL
N1267-23 - 1	2ND-3RD-FLOOR-WALLS-2	43	0.0121	-7470A	PPB	26 Jan 2022 15:17:31	U	SMPL
N1267-23DUP - 1	2ND-3RD-FLOOR-WALLS-2DUP	34	0.0099	-7470A	PPB	26 Jan 2022 15:19:48	U	SMPL
N1267-23MS - 1	2ND-3RD-FLOOR-WALLS-2MS	14693	3.6747	-7470A	PPB	26 Jan 2022 15:22:04	U	SMPL
N1267-23MSD - 1	2ND-3RD-FLOOR-WALLS-2MSD	14465	3.6177	-7470A	PPB	26 Jan 2022 15:24:21	U	SMPL
PB142291BL - 1	PBW	-24	-0.0046	-7470A	PPB	26 Jan 2022 15:26:38	U	SMPL
PB142291BS - 1	LCSW	14064	3.5175	-7470A	PPB	26 Jan 2022 15:28:54	U	SMPL
CCV74 - 1	CCV74	19032	4.7595	-7470A	PPB	26 Jan 2022 15:31:10	U	SMPL
CCB74 - 1	CCB74	-36	-0.0076	-7470A	PPB	26 Jan 2022 15:33:26	U	SMPL
N1257-03 - 1	CPIT-COMP	134	0.0349	-7470A	PPB	26 Jan 2022 15:35:41	U	SMPL
N1257-03DUP - 1	CPIT-COMPDUP	91	0.0241	-7470A	PPB	26 Jan 2022 15:37:56	U	SMPL
N1257-03MS - 1	CPIT-COMPMS	14650	3.6640	-7470A	PPB	26 Jan 2022 15:40:12	U	SMPL
N1257-03MSD - 1	CPIT-COMPMSD	14356	3.5905	-7470A	PPB	26 Jan 2022 15:45:09	U	SMPL
PB142264TB - 1	PB142264TB	-21	-0.0039	-7470A	PPB	26 Jan 2022 15:47:24	U	SMPL
N1267-23LX5 - 1		-8	-0.0006	-7470A	PPB	26 Jan 2022 15:49:40	U	SMPL
N1267-23A - 1		15859	3.9662	-7470A	PPB	26 Jan 2022 15:51:57	U	SMPL
N1257-03LX5 - 1		142	0.0369	-7470A	PPB	26 Jan 2022 15:56:49	U	SMPL
N1257-03A - 1		16558	4.1410	-7470A	PPB	26 Jan 2022 15:59:05	U	SMPL
CCV75 - 1	CCV75	18987	4.7483	-7470A	PPB	26 Jan 2022 16:01:21	U	SMPL
CCB75 - 1	CCB75	-31	-0.0064	-7470A	PPB	26 Jan 2022 16:03:37	U	SMPL